

© Copyright SEK. Reproduction in any form without permission is prohibited.

## Bärbara instrument för rökgasmätning – Del 2: Prestandafordringar på instrument för föreskriven emissionskontroll

*Specification for portable electrical apparatus designed to  
measure combustion flue gas parameters of heating appliances –  
Part 2: Performance requirements for apparatus used in  
statutory inspections and assessment*

Som svensk standard gäller europastandarden EN 50379-2:2012. Den svenska standarden innehåller den officiella engelska språkversionen av EN 50379-2:2012.

### Nationellt förord

Standarden ska användas tillsammans med SS-EN 50379-1.

Tidigare fastställd svensk standard SS-EN 50379-2, utgåva 1, 2004, gäller ej fr o m 2015-03-19.

---

ICS 13.320

---

Denna standard är fastställd av SEK Svensk Elstandard,  
som också kan lämna upplysningar om **sakinnehållet** i standarden.  
Postadress: SEK, Box 1284, 164 29 KISTA  
Telefon: 08 - 444 14 00. Telefax: 08 - 444 14 30  
E-post: sek@elstandard.se. Internet: www.elstandard.se

---

### *Standarder underlättar utvecklingen och höjer elsäkerheten*

Det finns många fördelar med att ha gemensamma tekniska regler för bl a säkerhet, prestanda, dokumentation, utförande och skötsel av elprodukter, elanläggningar och metoder. Genom att utforma sådana standarder blir säkerhetskraven tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

### *SEK är Sveriges röst i standardiseringsarbetet inom elområdet*

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

### *Stora delar av arbetet sker internationellt*

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

### *Var med och påverka!*

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

### **SEK Svensk Elstandard**

Box 1284  
164 29 Kista  
Tel 08-444 14 00  
[www.elstandard.se](http://www.elstandard.se)

English version

**Specification for portable electrical apparatus designed to measure  
combustion flue gas parameters of heating appliances -  
Part 2: Performance requirements for apparatus used in statutory  
inspections and assessment**

Spécification pour les appareils  
électriques portatifs conçus pour mesurer  
les paramètres des gaz de combustion  
dans les conduits d'évacuation des  
appareils de chauffage -  
Partie 2: Prescriptions des  
caractéristiques des appareils utilisés au  
cours des inspections et évaluations  
réglementaires

Anforderungen an tragbare elektrische  
Geräte zur Messung von  
Verbrennungsparametern von  
Heizungsanlagen -  
Teil 2: Anforderungen an das  
Betriebsverhalten von Geräten für den  
Einsatz bei gesetzlich geregelten  
Messungen und Beurteilungen

This European Standard was approved by CENELEC on 2012-03-19. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

**CENELEC**

European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**Management Centre: Avenue Marnix 17, B - 1000 Brussels**

## Contents

Foreword .....	3
Introduction .....	4
1 Scope .....	5
2 Normative references .....	5
3 Terms and definitions .....	5
4 General requirements .....	5
5 Test methods and performance requirements .....	6
5.1 General requirements for tests .....	6
5.2 Normal conditions for tests .....	6
5.3 Mechanical tests .....	6
5.4 Electrical and software tests .....	7
5.5 Tests with test gases .....	7
5.6 Tests with real flue gases .....	8
5.7 Calculated values .....	9
5.8 Temperature .....	9
5.9 Pressure .....	10
Bibliography .....	11

## Foreword

This document (EN 50379-2:2012) has been prepared by CENELEC TC 216 "Gas detectors".

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-03-19
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2015-03-19

This document supersedes EN 50379-2:2004.

EN 50379-2:2012 includes the following significant technical changes with respect to EN 50379-2:2004:

- 5.5.7 considers calibration curves for sensors with nonlinear signal;
- 5.5.8 considers influence of pressure variations;
- 5.5.9 considers the influence of water vapour on the gas signal;
- 5.7.2 for calculated values was amended;
- 5.7.4 considers 15 min. average concentration for solid fuels;
- 5.9.1 was amended to cover measurement at the circular orifice.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

---

## Introduction

This European Standard covers apparatus for measuring gas concentrations and other combustion parameters, as used in the installation and maintenance of heating appliances. It forms a specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances, and includes the following parts under the generic title *Specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances*:

Part 1: General requirements and test methods;

Part 2: Performance requirements for apparatus used in statutory inspections and assessments;

Part 3: Performance requirements for apparatus used in non-statutory servicing of gas fired heating appliances.

EN 50379-1 specifies general requirements for the construction, testing and performance of portable spot reading apparatus designed to give an assessment of specific combustion flue gas parameters such as concentration of gaseous compounds, temperature and/or pressure to check the combustion performance of heating appliances for domestic residential and commercial applications using commercially available fuels.

EN 50379-2 is for apparatus intended to be used for statutory measurement. In several European countries, legal requirements exist for the performance of heating appliances. Authorised inspectors use these apparatus to measure the flue gas parameters, in order to test compliance with national regulations. Due to the legal consequences resulting from the measurement there are strict requirements regarding the measuring uncertainty of these apparatus. Therefore EN 50379-2 includes maximum values for measuring uncertainty of the apparatus. Tests with real flue gases form a key part of the verification of the performance of the apparatus for statutory measurement. The measuring uncertainty has to be justified by internationally accepted methods over the whole measuring range.

EN 50379-3 is for apparatus intended to be used for non-statutory applications. There are reduced performance requirements, because the apparatus are designed to decide if maintenance for a gas fired appliance is required, and for adjusting the appliance during maintenance. There will be no determination of the measuring uncertainty for the apparatus.

## 1 Scope

This European Standard covers apparatus designed to measure flue gas parameters of heating appliances for domestic residential and commercial applications using commercially available fuels in compliance with metrological specification.

The apparatus may consist of different functional modules that may be tested separately for complying with this standard and will be combined in different ways according to the different applications. Part 1 of EN 50379 specifies the general requirements and is supplemented by the requirements in EN 50379-2 and/or EN 50379-3.

This European Standard specifies the performance requirements of portable spot reading apparatus designed to give a measurement of specific combustion flue gas parameters such as concentration of gaseous compounds, temperature and/or pressure to be used for testing the compliance with national regulations for the above mentioned appliances.

This European Standard excludes apparatus for

- continuous emission, safety monitoring and control, and
- use in vessels with an international load line.

NOTE 1 When this apparatus is used in industrial premises, national regulations should be observed.

NOTE 2 Apparatus may contain functional modules that are not covered by this standard e.g. measurement of smoke spot number (see EN 267:2009+A1:2011, Annex A) and/or measurement of indoor ambient air (see EN 50543).

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 50270:2006, *Electromagnetic compatibility – Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen*

EN 50271:2010, *Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen – Requirements and tests for apparatus using software and/or digital technologies*

EN 50379-1:2012, *Specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances – Part 1: General requirements and test methods*

EN 60335-1:2002, *Household and similar electrical appliances – Safety – Part 1: General requirements (IEC 60335-1:2001, mod.)*

EN 60529:1991, *Degrees of protection provided by enclosures (IP Code) (IEC 60529:1989)*