

© Copyright SEK. Reproduction in any form without permission is prohibited.

Elektroakustik – Mätmikrofoner – Del 8: Metoder för jämförande bestämning av frifältskänslighet för standardiserade mikrofoner

*Measurement microphones –
Part 8: Methods for determining the free-field sensitivity of
working standard microphones by comparison*

Som svensk standard gäller europastandarden EN 61094-8:2012. Den svenska standarden innehåller den officiella engelska språkversionen av EN 61094-8:2012.

Nationellt förord

Europastandarden EN 61094-8:2012^{*)}

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 61094-8, First edition, 2012 - Measurement microphones - Part 8: Methods for determining the free-field sensitivity of working standard microphones by comparison**

utarbetad inom International Electrotechnical Commission, IEC.

^{*)} Corrigendum, January 2013 till EN 61094-8:2012 är inarbetat i standarden.

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a säkerhet, prestanda, dokumentation, utförande och skötsel av elprodukter, elanläggningar och metoder. Genom att utforma sådana standarder blir säkerhetskraven tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284
164 29 Kista
Tel 08-444 14 00
www.elstandard.se

**Measurement microphones -
Part 8: Methods for determining the free-field sensitivity of working
standard microphones by comparison
(IEC 61094-8:2012)**

Microphones de mesure -
Partie 8: Méthodes pour la détermination
de l'efficacité en champ libre par
comparaison des microphones étalons de
travail
(CEI 61094-8:2012)

Messmikrofone -
Teil 8: Verfahren zur Ermittlung des
Freifeld-Übertragungskoeffizienten von
Gebrauchs-Normalmikrofonen nach der
Vergleichsmethode
(IEC 61094-8:2012)

This European Standard was approved by CENELEC on 2012-10-24. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 29/752/CDV, future edition 1 of IEC 61094-8, prepared by IEC/TC 29 "IEC TC 29, Electroacoustics" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61094-8:2012.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-07-24
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2015-10-24

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 61094-8:2012 was approved by CENELEC as a European Standard without any modification.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61094-1	-	Measurement microphones - Part 1: Specifications for laboratory standard microphones	EN 61094-1	-
IEC 61094-2	-	Electroacoustics - Measurement microphones - Part 2: Primary method for the pressure calibration of laboratory standard microphones by the reciprocity technique	EN 61094-2	-
IEC 61094-3	-	Measurement microphones - Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique	EN 61094-3	-
IEC 61094-4	-	Measurement microphones - Part 4: Specifications for working standard microphones	EN 61094-4	-
IEC 61094-5	-	Measurement microphones - Part 5: Methods for pressure calibration of working standard microphones by comparison	EN 61094-5	-
IEC 61094-6	-	Measurement microphones - Part 6: Electrostatic actuators for determination of frequency response	EN 61094-6	-
IEC/TS 61094-7	-	Measurement microphones - Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones	-	-
ISO/IEC Guide 98-3	-	Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)	-	-
ISO 26101	-	Acoustics - Test methods for the qualification of free-field environments	-	-

CONTENTS

1	Scope	6
2	Normative references	6
3	Terms and definitions	7
4	Reference environmental conditions	8
5	Principles of free-field calibration by comparison	8
5.1	General principle	8
5.2	General principles using sequential excitation	8
5.3	General principles using simultaneous excitation	8
6	General requirements	9
6.1	The test space	9
6.2	Methods of establishing the free-field	9
6.2.1	General	9
6.2.2	Using a test space with sound absorbing surfaces	9
6.2.3	Time selective methods for obtaining the free-field sensitivity	10
6.3	The sound source	10
6.4	Reference microphone	11
6.5	Monitor microphone	12
6.6	Test signals	12
6.7	Configuration for the reference microphone and microphone under test	13
7	Factors influencing the free-field sensitivity	13
7.1	General	13
7.2	Polarizing voltage	13
7.3	Acoustic centre of the microphone	13
7.4	Angle of incidence and alignment with the sound source	14
7.5	Mounting configuration	14
7.6	Dependence on environmental conditions	14
8	Calibration uncertainty components	14
8.1	General	14
8.2	Sensitivity of the reference microphone	15
8.3	Measurement of the microphone output	15
8.4	Differences between the sound pressure applied to the reference microphone and to the microphone under test	15
8.5	Influence of indirect sound	15
8.6	Influence of signal processing	16
8.7	Influence of microphone characteristics and measurement system performance	16
8.7.1	Microphone capacitance	16
8.7.2	Measurement system non-linearity	16
8.7.3	Validation of calibration system	16
8.8	Uncertainty on free-field sensitivity level	16
Annex A (informative)	Basic substitution calibration in a free-field chamber	18
Annex B (informative)	Time selective techniques	22
Bibliography		30

Figure A.1 – Illustration of source and receiver setup in a free-field room, where the monitor microphone has been integrated into the loudspeaker	18
Figure A.2 – Practical implementation in a hemi-anechoic room with a source flush-mounted in the floor	19
Figure A.3 – Examples of loudspeaker sources	21
Figure B.1 – Illustration of set-up for measurement with time selective techniques.....	23
Table 1 – Calibration options for the reference microphone and associated typical measurement uncertainty.....	12
Table 2 – Typical uncertainty components	17

MEASUREMENT MICROPHONES –

Part 8: Methods for determining the free-field sensitivity of working standard microphones by comparison

1 Scope

This part of the IEC 61094 series is applicable to working standard microphones meeting the requirements of IEC 61094-4. It describes methods of determining the free-field sensitivity by comparison with a laboratory standard microphone or working standard microphone (where applicable) that has been calibrated according to either:

- IEC 61094-3,
- IEC 61094-2 or IEC 61094-5, and where factors given in IEC/TS 61094-7 have been applied,
- IEC 61094-6,
- this part of IEC 61094.

Methods performed in an acoustical environment that is a good approximation to an ideal free-field (e.g. a high quality free-field chamber), and methods that use post processing of results to minimise the effect of imperfections in the acoustical environment, to simulate free-field conditions, are both covered by this part of IEC 61094. Comparison methods based on the principles described in IEC 61094-3 are also possible but beyond the scope of this part of IEC 61094.

NOTE 1 This part of IEC 61094 is also applicable to laboratory standard microphones meeting the requirements of IEC 61094-1, noting that these microphones also meet the electroacoustic specifications for working standard microphones.

NOTE 2 This part of IEC 61094 is also applicable to combinations of microphone and preamplifier where the determined sensitivity is referred to the unloaded output voltage of the preamplifier.

NOTE 3 Other devices, for example, sound level meters can be calibrated using the principles of this part of IEC 61094, but are not within the scope of this standard.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61094-1, *Measurement microphones – Part 1: Specifications for laboratory standard microphones*

IEC 61094-2, *Electroacoustics – Measurement microphones – Part 2: Primary method for pressure calibration of laboratory standard microphones by the reciprocity technique*

IEC 61094-3, *Measurement microphones – Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique*

IEC 61094-4, *Measurement microphones – Part 4: Specifications for working standard microphones*

IEC 61094-5, *Measurement microphones – Part 5: Methods for pressure calibration of working standard microphones by comparison*

IEC 61094-6, *Measurement microphones – Part 6: Electrostatic actuators for determination of frequency response*

IEC/TS 61094-7, *Measurement microphones – Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 26101, *Acoustics – Test methods for the qualification of free-field environments*