

© Copyright SEK. Reproduction in any form without permission is prohibited.

## **Elektrisk utrustning för medicinskt bruk – Säkerhet och väsentliga prestanda – Del 2-26: Särskilda fordringar på EEG-apparater**

*Medical electrical equipment –  
Part 2-26: Particular requirements for the basic safety and  
essential performance of electroencephalographs*

Som svensk standard gäller europastandarden EN 60601-2-26:2015. Den svenska standarden innehåller den officiella engelska språkversionen av EN 60601-2-26:2015.

### **Nationellt förord**

Europastandarden EN 60601-2-26:2015

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 60601-2-26, Third edition, 2012 - Medical electrical equipment - Part 2-26: Particular requirements for the basic safety and essential performance of electroencephalographs**

utarbetad inom International Electrotechnical Commission, IEC.

Standarden ska användas tillsammans med SS-EN 60601-1.

Tidigare fastställd svensk standard SS-EN 60601-2-26, utgåva 2, 2003, gäller ej fr o m 2018-04-14.

### *Standarder underlättar utvecklingen och höjer elsäkerheten*

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

### *SEK är Sveriges röst i standardiseringsarbetet inom elområdet*

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

### *Stora delar av arbetet sker internationellt*

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

### *Var med och påverka!*

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

### **SEK Svensk Elstandard**

Box 1284  
164 29 Kista  
Tel 08-444 14 00  
[www.elstandard.se](http://www.elstandard.se)

English Version

**Medical electrical equipment - Part 2-26: Particular requirements  
for the basic safety and essential performance of  
electroencephalographs  
(IEC 60601-2-26:2012)**

Appareils électromédicaux - Partie 2-26: Exigences  
particulières pour la sécurité de base et les performances  
essentielles des électroencéphalographes  
(IEC 60601-2-26:2012)

Medizinische elektrische Geräte - Teil 2-26: Besondere  
Festlegungen für die Sicherheit einschließlich der  
wesentlichen Leistungsmerkmale von  
Elektroenzephalographen  
(IEC 60601-2-26:2012)

This European Standard was approved by CENELEC on 2015-04-14. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## **Foreword**

The text of document 62D/990/FDIS, future edition 3 of IEC 60601-2-26, prepared by SC 62D "Electromedical equipment", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-26:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-01-14
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-04-14

This document supersedes EN 60601-2-26:2003.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document.

## **Endorsement notice**

The text of the International Standard IEC 60601-2-26:2012 was approved by CENELEC as a European Standard without any modification.

## Annex ZA

(normative)

### Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cenelec.eu](http://www.cenelec.eu).

***Annex ZA of EN 60601-1:2006 applies, except as follows:***

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<b><i>In Annex ZA of EN 60601-1:2006 replace IEC 60601-1-2 by:</i></b>				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment -	EN 60601-1-2	2007
-	-	Part 1-2: General requirements for basic safety and essential performance -	+ corrigendum Mar.	2010
		Collateral standard: Electromagnetic compatibility - Requirements and tests		

***Add to Annex ZA of EN 60601-1:2006 the following new references:***

IEC 60601-1	2005	Medical electrical equipment -	EN 60601-1	2006
-	-	Part 1: General requirements for basic safety and essential performance	+ corrigendum Mar.	2010
-	-		+ A12	2014
IEC 60601-2-27	2011	Medical electrical equipment -	EN 60601-2-27	2014
		Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment		

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
201.1 Scope, object and related standards .....	6
201.2 Normative references .....	8
201.3 Terms and definitions.....	8
201.4 General requirements.....	9
201.5 General requirements for testing of ME EQUIPMENT.....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS .....	10
201.7 ME EQUIPMENT identification, marking and documents.....	11
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	12
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	17
201.10 Protection against unwanted and excessive radiation HAZARDS.....	17
201.11 Protection against excessive temperatures and other HAZARDS.....	17
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	18
201.13 HAZARDOUS SITUATIONS and fault conditions .....	23
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	23
201.15 Construction of ME EQUIPMENT .....	23
201.16 ME SYSTEMS .....	23
201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS.....	23
202 ELECTROMAGNETIC COMPATIBILITY – Requirements and tests .....	23
Annexes .....	26
Annex AA (informative) Particular guidance and rationale .....	27
Index of defined terms used in this particular standard.....	29
Figure 201.101 – Test of protection against the effects of defibrillation (common mode) .....	14
Figure 201.102 – Test of protection against the effects of defibrillation (differential mode) .....	15
Figure 201.103 – Application of the test voltage between LEAD WIRES to test the energy delivered by the defibrillator.....	16
Figure 201.104 – General test circuit .....	20
Figure 201.105 – Test circuit for COMMON MODE REJECTION.....	22
Figure 202.101 – Test layout for radiated and conducted EMISSION test and radiated immunity test .....	24
Figure 202.102 – Set-up for radiated immunity test according to 202.6.2.3.2 .....	25
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	10

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –****Part 2-26: Particular requirements for the basic safety  
and essential performance of electroencephalographs**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-26 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition of IEC 60601-2-26 published in 2002. This edition constitutes a technical revision to the new structure of the third edition (2005) of IEC 60601-1.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/990/FDIS	62D/1012/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.



## INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROENCEPHALOGRAPHS. It amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*), hereinafter referred to as the general standard.

The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this standard.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-26: Particular requirements for the basic safety and essential performance of electroencephalographs

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1</sup> applies, except as follows:

##### 201.1.1 \* Scope

*Replacement:*

This particular standard applies to BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROENCEPHALOGRAPHS (EEG) as defined in 201.3.63, hereafter referred to as ME EQUIPMENT. This standard is applicable to ME EQUIPMENT used in a clinical environment (e.g., hospital, physician's office, etc.).

This standard does not cover requirements for other equipment used in electroencephalography such as:

- phono-photoc stimulators;
- electroencephalographic telemetry;
- EEG data storage and retrieval;
- ME EQUIPMENT particularly intended for monitoring during electro-convulsive therapy;
- ambulatory electroencephalographic recorders.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ELECTROENCEPHALOGRAPHS as defined in 201.3.63.

##### 201.1.3 Collateral standards

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

<sup>1</sup> The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3 and IEC 60601-1-10<sup>2</sup> do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

#### **201.1.4 Particular standards**

##### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

---

<sup>2</sup> IEC 60601-1-10:2007, *Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers*

## **201.2 Normative references**

Clause 2 of the general standard applies, except as follows:

*Amendment:*

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

*Addition:*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-2-27:2011, *Medical electrical equipment – Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*