

© Copyright SEK. Reproduction in any form without permission is prohibited.

Elektrisk utrustning för medicinskt bruk – Del 2-25: Särskilda fordringar på säkerhet och väsentliga prestanda för EKG-utrustning

*Medical electrical equipment –
Part 2-25: Particular requirements for the basic safety and
essential performance of electrocardiographs*

Som svensk standard gäller europastandarden EN 60601-2-25:2015. Den svenska standarden innehåller den officiella engelska språkversionen av EN 60601-2-25:2015.

Nationellt förord

Europastandarden EN 60601-2-25:2015

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 60601-2-25, Second edition, 2011 - Medical electrical equipment - Part 2-25: Particular requirements for the basic safety and essential performance of electrocardiographs**

utarbetad inom International Electrotechnical Commission, IEC.

Standarden ska användas tillsammans med SS-EN 60601-1.

Tidigare fastställd svensk standard SS-EN 60601-2-25, utgåva 1, 1996, SS-EN 60601-2-25/A1, utgåva 1, 1999 och SS-EN 60601-2-51, utg 1, 2003, gäller ej fr o m 2018-09-15.

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284
164 29 Kista
Tel 08-444 14 00
www.elstandard.se

English Version

**Medical electrical equipment - Part 2-25: Particular requirements
for the basic safety and essential performance of
electrocardiographs
(IEC 60601-2-25:2011)**

Appareils électromédicaux - Partie 2-25: Exigences
particulières pour la sécurité de base et les performances
essentielles des électrocardiographes
(IEC 60601-2-25:2011)

Medizinische elektrische Geräte - Teil 2-25: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Elektrokardiographen
(IEC 60601-2-25:2011)

This European Standard was approved by CENELEC on 2015-09-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

The text of document 62D/944/FDIS, future edition 2 of IEC 60601-2-25, prepared by SC 62D "Electromedical equipment", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-25:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-06-15
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-09-15

This document supersedes EN 60601-2-25:1995 and EN 60601-2-51:2003.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document.

Endorsement notice

The text of the International Standard IEC 60601-2-25:2011 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-27	NOTE	Harmonized as EN 60601-2-27.
IEC 60601-2-47	NOTE	Harmonized as EN 60601-2-47.

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement in Annex ZA of EN 60601-1:2006:</i>				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment -	EN 60601-1-2	2007
-	-	Part 1-2: General requirements for basic safety and essential performance -	+ corrigendum Mar.	2010
		Collateral standard: Electromagnetic compatibility - Requirements and tests		
<i>Addition to Annex ZA of EN 60601-1:2006:</i>				
IEC 60601-2-2	2009	Medical electrical equipment -	EN 60601-2-2	2009
-	-	Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	+ A11	2011

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
201.1 Scope, object and related standards.....	8
201.2 Normative references	10
201.3 Terms and definitions	10
201.4 General requirements.....	12
201.5 General requirements for testing of ME EQUIPMENT	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	13
201.7 ME EQUIPMENT identification, marking and documents	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	16
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	21
201.10 Protection against unwanted and excessive radiation HAZARDS	21
201.11 Protection against excessive temperatures and other HAZARDS.....	21
201.12 Accuracy of controls and instruments and protection against hazardous outputs	22
201.13 HAZARDOUS SITUATIONS and fault conditions	37
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	37
201.15 Construction of ME EQUIPMENT	37
201.16 ME SYSTEMS.....	37
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	37
202 Electromagnetic compatibility – Requirements and tests	38
Annexes	43
Annex AA (informative) Particular guidance and rationale	44
Annex BB (informative) ELECTRODES, their positions, identifications and colour codes	51
Annex CC (informative) LEADS, their identification and colour codes (other than those specified in 201.12.4.102).....	53
Annex DD (informative) Polarity of PATIENT LEADS (other than those specified in 201.12.4.102)	54
Annex EE (informative) Additional marking of ELECTRODES.....	55
Annex FF (informative) Definitions and rules for the measurement of ELECTROCARDIOGRAMS	56
Annex GG (informative) Calibration and test data sets	61
Annex HH (informative) CTS test atlas.....	63
Bibliography.....	94
Index of defined terms used in this particular standard.....	95
Figure 201.101 – ELECTRODE position according to Frank	14
Figure 201.102 – Test of protection against the effects of defibrillation (differential mode) (see 201.8.5.5.1).....	19
Figure 201.103 – Test of protection against the effects of defibrillation (common mode) (see 201.8.5.5.1)	20
Figure 201.104 – Application of the test voltage between LEAD WIRES to test the energy delivered by the defibrillator.....	21

Figure 201.105 – Test circuit for COMMON MODE REJECTION and NOISE level	28
Figure 201.106 – General test circuit	30
Figure 201.107 – Triangular waveforms for test E of Table 201.107	32
Figure 201.108 – Input impulse signal and ELECTROCARDIOGRAPH response	32
Figure 201.109 – Circuit for test of linearity.....	34
Figure 201.110 – Result of linearity test.....	34
Figure 201.111 – Pacemaker overload test circuit.....	36
Figure 202.101 – Set-up for radiated and conducted emission test	39
Figure 202.102 – Set-up for radiated immunity test	40
Figure 202.103 – Test circuit for HF surgery protection measurement.....	42
Figure 202.104 – Test setup for HF surgery protection measurement.....	43
Figure BB.1a – LEADS and colours for fetal ECG (see Table BB.2)	52
Figure BB.1b – Positions of the ELECTRODES on the fetus for fetal ECG (see Table BB.2)	52
Figure BB.2 – LEAD positions and colours for fetal scalp ECG (see Table BB.2)	52
Figure FF.1 – Normal ELECTROCARDIOGRAM.....	56
Figure FF.2 – Determination of global intervals (example)	57
Figure FF.3 – Waveform durations, isoelectric segments	58
Figure FF.4 – QRS complex with small R-wave(s) (see Figure FF.5, FF.6)	59
Figure FF.5 – Detail of small accepted R-wave	60
Figure FF.6 – Detail of small rejected R-wave.....	60
Figure HH.1 – Nomenclature of calibration ECGS	66
Figure HH.2 – Nomenclature of analytical ECGS	69
Table 201.101 – ESSENTIAL PERFORMANCE requirements	12
Table 201.102 – ELECTRODES, their position, identification and colour code	14
Table 201.103 – Protection against the effect of defibrillation (test conditions).....	18
Table 201.104 – Acceptable mean differences and standard deviations for global intervals and Q-, R-, S-durations on calibration and analytical ECGS	23
Table 201.105 – Acceptable mean differences and standard deviations for global durations and intervals for biological ECGS	23
Table 201.106 – LEADS and their identification (nomenclature and definition).....	25
Table 201.107 – Frequency response	31
Table 201.108 – PATIENT ELECTRODE connection for pacemaker pulse display test	37
Table AA.1 – ELECTRODE positions and electrical strength requirements	46
Table BB.1 – ELECTRODES, their positions, identifications and colour codes (other than described in 201.7.4.101, Table 201.106)	51
Table BB.2 – Other ELECTRODE-positions, identifications and colour codes not covered by this particular standard.....	51
Table DD.1 – ELECTRODE polarities	54
Table EE.1 – Recommended identification and colour code for a 14-wire PATIENT CABLE	55
Table GG.1 – CALIBRATION and analytical ECGS	61
Table GG.2 – Data set for testing of measurement and wave recognition accuracy of biological data – 100 selected ECGS of the CSE-study with their numbering in the CSE database, to be used in 201.12.1.101.3.2	62
Table HH.1 – Naming of signals (calibration ECGS)	67

Table HH.2 – Naming of signals (analytical ECGs)	68
--	----

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-25: Particular requirements for the basic safety
and essential performance of electrocardiographs**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-25 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition of IEC 60601-2-25, published in 1993 and the first edition of IEC 60601-2-51, published in 2003. This second edition of IEC 60601-2-25 constitutes a technical revision of both those standards.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/944/FDIS	62D/957/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROCARDIOGRAPHIC EQUIPMENT. It amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

This particular standard now includes the contents of the particular standard IEC 60601-2-51: *Medical electrical equipment – Part 2-51: Particular requirements for the safety, including essential performance, of recording and analysing single channel and multichannel electrocardiographs*.

Updating the particular standards to refer to the third edition of the general standard provided the opportunity to merge the first editions of IEC 60601-2-25 and IEC 60601-2-51 into one standard. Reformatting and technical changes were both made.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. Knowledge of the reasons for these requirements will not only facilitate proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-25: Particular requirements for the basic safety and essential performance of electrocardiographs

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This particular standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROCARDIOGRAPHS as defined in 201.3.63 intended by themselves or as a part of an ME SYSTEM, for the production of ECG REPORTS for diagnostic purposes, hereinafter referred to as ME EQUIPMENT.

Not included within the scope of this particular standard are:

- a) the part of ME EQUIPMENT that provides vectorcardiographic loops;
- b) ambulatory electrocardiographic ME EQUIPMENT covered by IEC 60601-2-47 where not intended for obtaining ECG REPORTS for diagnostic purposes;
- c) cardiac monitors covered by IEC 60601-2-27 where not intended for obtaining ECG REPORTS for diagnostic purposes.

NOTE 1 For example. ME EQUIPMENT includes:

- a) direct-writing ELECTROCARDIOGRAPHS;
- b) other ME EQUIPMENT that produce ECG REPORTS for diagnostic purposes, e.g. patient monitors, defibrillators, exercise testing devices;
- c) ELECTROCARDIOGRAPHS having a display that is remote from the PATIENT (e.g. via phone lines, networks or storage media). These ME EQUIPMENT or ME SYSTEMS are within the scope of this particular standard excluding transmission media.

NOTE 2 ME EQUIPMENT that provide selection between diagnostic and monitoring functions shall meet the requirements of the appropriate standard when configured for that function.

ME EQUIPMENT intended for use under extreme or uncontrolled environmental conditions outside the hospital environment or physician's office, such as in ambulances and air transport, shall comply with this particular standard. Additional standards may apply to ME EQUIPMENT for those environments of use.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular requirements for BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROCARDIOGRAPHS as defined in 201.3.63.

201.1.3 Collateral standards

Addition:

¹ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3, IEC 60601-1-8 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

“Replacement” means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

“Addition” means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

“Amendment” means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, etc.

The term “this standard” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography beginning on page 94.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

Addition:

IEC 60601-2-2:2009, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*