



Fastställd 2017-06-14 Utgåva 3 Sida

1 (1+32)

Ansvarig kommitté

SFK TK 66

© Copyright SEK. Reproduction in any form without permission is prohibited.

# Elektrisk utrustning för mätning, styrning och för laboratorieändamål – Säkerhet –

Del 2-020: Särskilda fordringar på laboratoriecentrifuger

Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-020: Particular requirements for laboratory centrifuges

Som svensk standard gäller europastandarden EN 61010-2-020:2017. Den svenska standarden innehåller den officiella engelska språkversionen av EN 61010-2-020:2017.

# Nationellt förord

Europastandarden EN 61010-2-020:2017

består av:

- europastandardens ikraftsättningsdokument, utarbetat inom CENELEC
- IEC 61010-2-020, Third edition, 2016 Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges

utarbetad inom International Electrotechnical Commission, IEC.

Standarden ska användas tillsammans med SS-EN 61010-1, utgåva 3, 2010.

Tidigare fastställd svensk standard SS-EN 61010-2-020, utgåva 2, 2006, gäller ej fr o m 2020-04-28.

ICS 19.080.00; 71.040.10

### Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

# SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

### Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

### Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284 164 29 Kista Tel 08-444 14 00 www.elstandard.se

# EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 61010-2-020

April 2017

ICS 19.080; 71.040.10

Supersedes EN 61010-2-020:2006

## **English Version**

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges
(IEC 61010-2-020:2016)

Règles de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire - Partie 2-020: Exigences particulières pour centrifugeuses de laboratoire (IEC 61010-2-020:2016) Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte - Teil 2-020: Besondere Anforderungen an Laborzentrifugen (IEC 61010-2-020:2016)

This European Standard was approved by CENELEC on 2016-07-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

© 2017 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC Members.

# **European foreword**

The text of document 66/542/CDV, future edition 3 of IEC 61010-2-020, prepared by IEC/TC 66 "Safety of measuring, control and laboratory equipment" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61010-2-020:2017.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with (dow) 2020-04-28 the document have to be withdrawn

This document supersedes EN 61010-2-020:2006.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association.

### **Endorsement notice**

The text of the International Standard IEC 61010-2-020:2016 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standards indicated:

### Addition

IEC 60034 (series) NOTE Harmonized as EN 60034 (series).

# Annex ZA (normative)

# Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

# **Addition**

PublicationYearTitleEN/HDYearISO 3864seriesGraphical symbols - Safety colours andseries

safety signs - Part 1: Design principles for

safety signs and safety markings

# CONTENTS

FOR	FOREWORD3			
1	Scope and object	5		
2	Normative references	6		
3	Terms and definitions	6		
4	Tests	8		
5	Marking and documentation	8		
6	Protection against electric shock	11		
7	Protection against mechanical HAZARDS	11		
8	Mechanical resistance to shock and impact	17		
9	Protection against the spread of fire	17		
10	Equipment temperature limits and resistance to heat	17		
11	Protection against HAZARDS from fluids	17		
12	Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	19		
13	Protection against liberated gases and substances, explosion and implosion	19		
14	Components	19		
15	Protection by interlocks	19		
16	Hazards resulting from application	19		
17	Risk assessment	19		
Ann	Annexes20			
Annex L Index of defined terms				
Annex AA (normative) Dynamic microbiological test method for BIOSEALS2		21		
Ann	Annex BB (informative) General guidance and rationale for particular subclauses24			
	ex CC (informative) General guidance for an empirical method to determine the tic energy of a ROTOR	28		
Bibliography30				
Figu	Figure 101 – Rotor test setup28			
Tabl	Table 101 – Time-temperature conditions18			

# INTERNATIONAL ELECTROTECHNICAL COMMISSION

# SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

# Part 2-020: Particular requirements for LABORATORY CENTRIFUGES

### **FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-020 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

This third edition cancels and replaces the second edition published in 2006. It constitutes a technical revision and includes the following significant changes from the second edition:

- a) This Part 2 is established on the basis of the third edition (2010) of IEC 61010-1. The changes listed in its foreword affect this Part 2, too.
- b) The language has been updated to reflect current terminology for LABORATORY CENTRIFUGES used in the industry today.

It has the status of a group safety publication in accordance with IEC Guide 104.

The text of this standard is based on the following documents:

CDV	Report on voting
66/542/CDV	66/565A/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

This Part 2-020 is intended to be used in conjunction with IEC 61010-1. It was established on the basis of the third edition (2010).

This Part 2-020 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: Safety requirements for LABORATORY CENTRIFUGES.

Where a particular subclause of Part 1 is not mentioned in this Part 2, that subclause applies as far as is reasonable. Where this part states "addition", "modification" or "replacement", the relevant requirement, test specification or note in Part 1 should be adapted accordingly.

In this standard:

- 1) the following print types are used:
  - requirements: in roman type;
  - NOTES: in small roman type;
  - conformity and tests: in italic type;
  - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS.
- 2) subclauses, tables or figures which are additional to those in Part 1 are numbered starting from 101; additional annexes are lettered AA, BB, etc.

A list of all parts of the IEC 61010 series, under the general title: Safety requirements for electrical equipment for measurement, control, and laboratory use, may be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- · reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- · amended.

# SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

# Part 2-020: Particular requirements for laboratory centrifuges

# 1 Scope and object

This clause of Part 1 is applicable except as follows:

### 1.1.1 Scope

Replacement:

This Part 2 is applicable to electrically powered LABORATORY CENTRIFUGES.

This group safety publication is primarily intended to be used as a product safety standard for the products mentioned in the scope, but shall also be used by technical committees in the preparation of its publications for products similar to those mentioned in the scope of this standard, in accordance with the principles laid down in IEC Guide 104 and ISO/IEC Guide 51.

NOTE If all or part of the equipment falls within the scope of one or more other Part 2 standards of IEC 61010 as well as within the scope of this standard, it will also need to meet the requirements of those other Part 2 standards.

### 1.1.2 Equipment excluded from scope

Addition:

Add the following new item:

aa) IEC 60034 (Rotating electrical machinery);

# 1.2 Object

# 1.2.1 Aspects included in scope

Addition:

Add the following new items:

- aa) contact with moving parts (see 7.3);
- bb) LABORATORY CENTRIFUGE movement during any DISRUPTION (see 7.3.101);
- cc) high energy chemical reaction after ROTOR DISRUPTION (see 7.7.2.2 l));
- dd) ineffectiveness of BIOSEALS (see 13.101)

# 1.2.2 Aspects excluded from scope

Addition:

Add the following new items:

aa) additional precautions which may need to be observed when centrifuging materials which are flammable or explosive (see 5.4.101);

bb) additional precautions which may need to be observed when centrifuging materials that could react chemically with sufficient vigour to cause a HAZARD (see 5.4.101).

# 1.4 Environmental conditions

# 1.4.1 Normal environmental conditions

Replacement:

Replace item c) by the following:

c) temperature 2 °C to 40 °C;

# 1.4.2 Extended environmental conditions

Replacement:

Replace item c) by the following:

c) ambient temperatures below 2 °C or above 40 °C;

### 2 Normative references

This clause of Part 1 is applicable except as follows:

Addition:

ISO 3864 (all parts), Graphical symbols – Safety colours and safety signs