

© Copyright SEK. Reproduction in any form without permission is prohibited.

Mätning av elkvalitet i elnät – Del 2: Funktionsprovning och fordringar beträffande osäkerhet

*Power quality measurement in power supply systems –
Part 2: Functional tests and uncertainty requirements*

Som svensk standard gäller europastandarden EN 62586-2:2017. Den svenska standarden innehåller den officiella engelska språkversionen av EN 62586-2:2017.

Nationellt förord

Europastandarden EN 62586-2:2017

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 62586-2, Second edition, 2017 - Power quality measurement in power supply systems - Part 2: Functional tests and uncertainty requirements**

utarbetad inom International Electrotechnical Commission, IEC.

Tidigare fastställd svensk standard SS-EN 62586-2, utgåva 1, 2014, gäller ej fr o m 2020-08-25.

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284
164 29 Kista
Tel 08-444 14 00
www.elstandard.se

English Version

**Power quality measurement in power supply systems -
Part 2: Functional tests and uncertainty requirements
(IEC 62586-2:2017)**

Mesure de la qualité de l'alimentation dans les réseaux
d'alimentation - Partie 2: Essais fonctionnels et exigences
d'incertitude
(IEC 62586-2:2017)

Messung der Spannungsqualität in
Energieversorgungssystemen - Teil 2: Funktionsprüfungen
und Anforderungen an die Messunsicherheit
(IEC 62586-2:2017)

This European Standard was approved by CENELEC on 2017-04-11. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

The text of document 85/525/CDV, future edition 2 of IEC 62586-2, prepared by IEC/TC 85 "Measuring equipment for electrical and electromagnetic quantities" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 62586-2:2017.

The following dates are fixed:

- latest date by which the document has to be (dop) 2018-02-25
implemented at national level by
publication of an identical national
standard or by endorsement
- latest date by which the national (dow) 2020-08-25
standards conflicting with the
document have to be withdrawn

This document supersedes EN 62586-2:2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 62586-2:2017 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60359	NOTE	Harmonized as EN 60359.
IEC 61000-4-30:2008	NOTE	Harmonized as EN 61000-4-30:2009 ¹⁾ (not modified).

¹⁾ Superseded by EN 61000-4-30:2015 (IEC 61000-4-30:2015): DOW = 2018-03-27.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61000-2-4	-	Electromagnetic compatibility (EMC) - Part 2-4: Environment - Compatibility levels in industrial plants for low-frequency conducted disturbances	EN 61000-2-4	-
IEC 61000-4-7	-	Electromagnetic compatibility (EMC) - Part 4-7: Testing and measurement techniques - General guide on harmonics and interharmonics measurements and instrumentation, for power supply systems and equipment connected thereto	EN 61000-4-7	-
IEC 61000-4-15	-	Electromagnetic compatibility (EMC) - Part 4-15: Testing and measurement techniques - Flickermeter - Functional and design specifications	EN 61000-4-15	-
IEC 61000-4-30	2015	Electromagnetic Compatibility (EMC) - Part 4-30: Testing and measurement techniques - Power quality measurement methods	EN 61000-4-30	2015
IEC 62586-1	2013	Power quality measurement in power supply systems - Part 1: Power Quality Instruments (PQI)	EN 62586-1	2014
ISO/IEC Guide 98-3 2008		Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)	-	-

CONTENTS

FOREWORD	9
INTRODUCTION	11
1 Scope	12
2 Normative references	12
3 Terms, definitions, abbreviated terms, notations and symbols	13
3.1 General terms and definitions	13
3.2 Terms and definitions related to uncertainty	13
3.3 Notations	14
3.3.1 Functions	14
3.3.2 Symbols and abbreviated terms	14
3.3.3 Indices	14
4 Requirements	14
4.1 Requirements for products complying with class A	14
4.2 Requirements for products complying with class S	15
5 Functional type tests common requirements	17
5.1 General philosophy for testing	17
5.1.1 System topology	17
5.1.2 Stabilization time	17
5.1.3 Measuring ranges	17
5.1.4 Single "power-system influence quantities"	19
5.1.5 "External influence quantities"	21
5.1.6 Test criteria	21
5.2 Testing procedure	22
5.2.1 Device under test	22
5.2.2 Testing conditions	22
5.2.3 Testing equipment	22
6 Functional testing procedure for instruments complying with class A according to IEC 61000-4-30	22
6.1 Power frequency	22
6.1.1 General	22
6.1.2 Measurement method	23
6.1.3 Measurement uncertainty and measuring range	23
6.1.4 Measurement evaluation	24
6.1.5 Measurement aggregation	24
6.2 Magnitude of supply voltage	24
6.2.1 Measurement method	24
6.2.2 Measurement uncertainty and measuring range	24
6.2.3 Measurement evaluation	25
6.2.4 Measurement aggregation	25
6.3 Flicker	27
6.4 Supply voltage interruptions, dips and swells	27
6.4.1 General	27
6.4.2 Check dips / interruptions in polyphase system	35
6.4.3 Check swells in polyphase system	37
6.5 Supply voltage unbalance	38
6.5.1 General	38

6.5.2	Measurement method, measurement uncertainty and measuring range	39
6.5.3	Aggregation	39
6.6	Voltage harmonics	39
6.6.1	Measurement method	39
6.6.2	Measurement uncertainty and measuring range	40
6.6.3	Measurement evaluation	41
6.6.4	Measurement aggregation	41
6.7	Voltage interharmonics	43
6.7.1	Measurement method	43
6.7.2	Measurement uncertainty and measuring range	44
6.7.3	Measurement evaluation	45
6.7.4	Measurement aggregation	45
6.8	Mains signalling voltages on the supply voltage	47
6.8.1	Measurement method	47
6.8.2	Measurement uncertainty and measuring range	49
6.8.3	Aggregation	50
6.9	Measurement of underdeviation and overdeviation parameters	50
6.9.1	Measurement method	50
6.9.2	Measurement uncertainty and measuring range	52
6.9.3	Measurement evaluation	53
6.9.4	Measurement aggregation	53
6.10	Flagging	56
6.11	Clock uncertainty testing	58
6.12	Variations due to external influence quantities	58
6.12.1	General	58
6.12.2	Influence of temperature	59
6.12.3	Influence of power supply voltage	61
6.13	Rapid voltage changes (RVC)	62
6.13.1	RVC parameters and evaluation	62
6.13.2	General	62
6.13.3	"No RVC" tests	64
6.13.4	"RVC threshold and setup" test	68
6.13.5	"RVC parameters" test	70
6.13.6	"RVC polyphase" tests	72
6.13.7	"Voltage is in steady-state condition" tests	74
6.14	Magnitude of current	77
6.15	Harmonic current	77
6.16	Interharmonic currents	77
6.17	Current unbalance	77
6.17.1	General	77
6.17.2	Measurement method, measurement uncertainty and measuring range	78
7	Functional testing procedure for instruments complying with class S according to IEC 61000-4-30	78
7.1	Power frequency	78
7.1.1	General	78
7.1.2	Measurement method	79
7.1.3	Measurement uncertainty and measuring range	79
7.1.4	Measurement evaluation	80
7.1.5	Measurement aggregation	80

7.2	Magnitude of the supply voltage	80
7.2.1	Measurement method	80
7.2.2	Measurement uncertainty and measuring range	80
7.2.3	Measurement evaluation	81
7.2.4	Measurement aggregation	81
7.3	Flicker.....	83
7.4	Supply voltage interruptions, dips and swells	83
7.4.1	General requirements	84
7.4.2	Check dips / interruptions in polyphase system.....	89
7.4.3	Check swells in polyphase system	91
7.5	Supply voltage unbalance	92
7.5.1	General	92
7.5.2	Measurement method, measurement uncertainty and measuring range	93
7.5.3	Aggregation	93
7.6	Voltage harmonics	93
7.6.1	General	93
7.6.2	Measurement method	94
7.6.3	Measurement method, measurement uncertainty and measuring range	95
7.6.4	Measurement evaluation.....	96
7.6.5	Measurement aggregation	96
7.7	Voltage interharmonics	98
7.8	Mains signalling voltages on the supply voltage	98
7.8.1	General	98
7.8.2	Measurement method	99
7.8.3	Measurement uncertainty and measuring range.....	99
7.8.4	Aggregation	99
7.9	Measurement of underdeviation and overdeviation parameters	99
7.10	Flagging.....	99
7.11	Clock uncertainty testing.....	101
7.12	Variations due to external influence quantities	102
7.12.1	General	102
7.12.2	Influence of temperature	103
7.12.3	Influence of power supply voltage	105
7.13	Rapid voltage changes.....	106
7.14	Magnitude of current.....	106
7.15	Harmonic current	106
7.16	Interharmonic currents	106
7.17	Current unbalance	106
7.17.1	General	106
7.17.2	Measurement method, measurement uncertainty and measuring range	107
8	Calculation of measurement uncertainty and operating uncertainty	108
Annex A (normative)	Intrinsic uncertainty and operating uncertainty,	110
A.1	General.....	110
A.2	Measurement uncertainty	110
A.3	Operating uncertainty.....	111
Annex B (informative)	Overall system uncertainty	112
Annex C (normative)	Calculation of measurement and operating uncertainty for voltage magnitude and power frequency	113

C.1	Selection of test points to verify operating uncertainty and uncertainty under reference conditions	113
C.2	Class A calculation examples.....	113
C.2.1	General	113
C.2.2	Parameter: magnitude of supply voltage, $U_{din} = 230\text{ V}$, 50/60Hz, rated range of temperature -25 °C to $+55\text{ °C}$	113
C.2.3	Parameter: power frequency 50/60 Hz, rated range of temperature -25 °C to $+55\text{ °C}$	114
Annex D (informative)	Further test on dips (amplitude and phase angles changes)	116
D.1	Phase-to-phase or phase-to-neutral testing.....	116
D.2	Test method.....	116
Annex E (informative)	Further tests on dips (polyphase): test procedure	118
E.1	General.....	118
E.2	Phase voltage dips and interruptions	119
E.3	Phase swells.....	119
Annex F (normative)	Gapless measurements of voltage amplitude and harmonics test	121
F.1	Purpose of the test.....	121
F.2	Test set up.....	121
F.3	Voltage amplitude	121
F.3.1	Test signal.....	121
F.3.2	Result evaluation	121
F.4	Harmonics	122
F.4.1	Test signal.....	122
F.4.2	Result evaluation	122
F.5	Inter-harmonics.....	123
F.5.1	Test signal.....	123
F.5.2	Result evaluation	123
Annex G (informative)	Gapless measurements of voltage amplitude and harmonics.....	124
Annex H (informative)	Testing equipment recommendations	133
H.1	Testing range.....	133
H.2	Uncertainty and stability of source and reference meter	133
H.2.1	Uncertainty of source and reference meter	133
H.2.2	Stability of the source	134
H.3	Time synchronisation	134
H.4	Power quality functions of source and reference meter	134
H.5	Traceability	135
Annex I (informative)	Recommendations related to a declaration of conformity (DoC) and a test report	136
I.1	Definitions.....	136
I.2	Recommendations	136
I.3	Example of IEC 62586-1 declaration of conformity	136
I.4	Example of IEC 62586-2 declaration	138
I.4.1	General	138
I.4.2	Recommendation for IEC 62586-2 test report	139
I.4.3	Recommendation for IEC 62586-2 test summary	140
I.4.4	Recommendation for IEC 62586-2 test equipment information	140
I.4.5	Recommendation for IEC 62586-2 tested functions.....	140
Bibliography		141

Figure 1 – Overview of test for dips according to test A4.1.1	30
Figure 2 – Detail 1 of waveform for test of dips according to test A4.1.1	31
Figure 3 – Detail 2 of waveform for tests of dips according to A4.1.1	31
Figure 4 – Detail 3 of waveform for tests of dips according to test A4.1.1.....	32
Figure 5 – Detail 1 of waveform for test of dips according to test A4.1.2	32
Figure 6 – Detail 2 of waveform for tests of dips according to test A4.1.2.....	33
Figure 7 – Detail 1 of waveform for test of swells according to test A4.1.2	33
Figure 8 – Detail 2 of waveform for tests of swells according to test A4.1.2	34
Figure 9 – Sliding reference voltage test.....	34
Figure 10 – Sliding reference start up condition	35
Figure 11 – Detail 1 of waveform for test of polyphase dips/interruptions	36
Figure 12 – Detail 2 of waveform for test of polyphase dips/interruptions	36
Figure 13 – Detail 3 of waveform for test of polyphase dips/interruptions	37
Figure 14 – Detail 1 of waveform for test of polyphase swells	38
Figure 15 – Detail 2 of waveform for test of polyphase swells	38
Figure 16 – Flagging test for class A.....	57
Figure 17 – Clock uncertainty testing	58
Figure 18 – Example of RVC event	62
Figure 19 – A13.1.1 waveform	65
Figure 20 – A13.1.1 waveform with RVC limits and arithmetic mean	65
Figure 21 – A13.1.2 waveform	66
Figure 22 – A13.1.2 waveform with RVC limits and arithmetic means.....	67
Figure 23 – A13.1.3 waveform	68
Figure 24 – A13.1.3 waveform with RVC limits and arithmetic mean	68
Figure 25 – A13.2.1 waveform	69
Figure 26 – A13.2.1 waveform with RVC limits and arithmetic mean	70
Figure 27 – A13.3.1 waveform	71
Figure 28 – A13.3.1 waveform with RVC limits and arithmetic mean	72
Figure 29 – A13.4.1 waveform	73
Figure 30 – A13.5.1 waveform	75
Figure 31 – A13.5.1 waveform with RVC limits and arithmetic mean	75
Figure 32 – A13.5.2 waveform	76
Figure 33 – A13.5.2 waveform with RVC limits and arithmetic mean	77
Figure 34 – Detail 1 of waveform for test of dips according to test S4.1.2	86
Figure 35 – Detail 2 of waveform for tests of dips according to test S4.1.2.....	87
Figure 36 – Detail 1 of waveform for test of swells according to test S4.1.2	87
Figure 37 – Detail 2 of waveform for tests of swells according to test S4.1.2.....	88
Figure 38 – Sliding reference voltage test.....	88
Figure 39 – Sliding reference start-up condition	89
Figure 40 – Detail 1 of waveform for test of polyphase dips/interruptions	90
Figure 41 – Detail 2 of waveform for test of polyphase dips/interruptions	90
Figure 42 – Detail 3 of waveform for test of polyphase dips/interruptions	91

Figure 43 – Detail 1 of waveform for test of polyphase swells	92
Figure 44 – Detail 2 of waveform for test of polyphase swells	92
Figure 45 – Flagging test for class S	101
Figure 46 – Clock uncertainty testing	102
Figure A.1 – Different kinds of uncertainties	110
Figure D.1 – Phase-to-neutral testing on three-phase systems	116
Figure D.2 – Phase-to-phase testing on three-phase systems	116
Figure E.1 – Example for one phase of a typical N cycle injection	118
Figure E.2 – Dip/interruption accuracy (amplitude and timing) test	119
Figure E.3 – Swell accuracy (amplitude and timing) test	120
Figure G.1 – Simulated signal under noisy conditions	124
Figure G.2 – Waveform for checking gapless RMS voltage measurement	125
Figure G.3 – 2,3 Hz frequency fluctuation	125
Figure G.4 – Spectral leakage effects for a missing sample	126
Figure G.5 – Illustration of Q_{RMS} for missing samples	127
Figure G.6 – Detection of a single missing sample	127
Figure G.7 – Q_{RMS} for an ideal signal, sampling error = -300×10^{-6}	128
Figure G.8 – Q_{RMS} for an ideal signal, sampling error = 400×10^{-6}	128
Figure G.9 – Q_{RMS} for an ideal signal, sampling error = 200×10^{-6}	129
Figure G.10 – $Q_H(5)$ with ideal test signal and perfect sampling frequency synchronization	130
Figure G.11 – $Q_H(5)$ with 300×10^{-6} sampling frequency error and 100×10^{-6} modulation frequency error	130
Figure G.12 – Q_{RMS} with a 20/24-cycle sliding window with an output every 10/12 cycles	131
Figure G.13 – Amplitude test for fluctuating component	131
Table 1 – Summary of type tests for class A	15
Table 2 – Summary of type tests for class S	16
Table 3 – Testing points for each measured parameter	18
Table 4 – List of single "power-system influence quantities"	20
Table 5 – Influence of temperature	21
Table 6 – Influence of auxiliary power supply voltage	21
Table 7 – List of generic test criteria	22
Table 8 – Specification of test A13.1.1	64
Table 9 – Specification of test A13.1.2	66
Table 10 – Specification of test A13.1.3	67
Table 11 – Specification of test A13.2.1	69
Table 12 – Specification of test A13.3.1	71
Table 13 – Specification of test A13.4.1	73
Table 14 – Specification of test A13.5.1	74
Table 15 – Specification of test A13.5.2	76
Table 16 – Uncertainty requirements	109
Table D.1 – Tests pattern	117

Table H.1 – Testing range.....	133
Table H.2 – Uncertainty of source and reference meter.....	134
Table H.3 – Stability of source	134
Table I.1 – Example of a DoC related to compliance with IEC 62586-1	137
Table I.2 – Example of DoC related to compliance with IEC 62586-2	139

INTERNATIONAL ELECTROTECHNICAL COMMISSION

POWER QUALITY MEASUREMENT IN POWER SUPPLY SYSTEMS –**Part 2: Functional tests and uncertainty requirements****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62586-2 has been prepared by IEC technical committee 85: Measuring equipment for electrical and electromagnetic quantities.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) test procedures for RVC and current have been added;
- b) mistakes have been fixed.

This bilingual version (2017-11) corresponds to the monolingual English version, published in 2017-03.

The text of this standard is based on the following documents:

CDV	Report on voting
85/525/CDV	85/571/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

The French version of this standard has not been voted upon.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 62586 series, published under the general title *Power quality measurement in power supply systems*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

Power quality is more and more important worldwide in power supply systems and is generally assessed by power quality instruments.

This part of IEC 62586 specifies functional and uncertainty tests intended to verify the compliance of a product to class A and class S measurement methods defined in IEC 61000-4-30.

This document therefore complements IEC 61000-4-30.

POWER QUALITY MEASUREMENT IN POWER SUPPLY SYSTEMS –

Part 2: Functional tests and uncertainty requirements

1 Scope

This part of IEC 62586 specifies functional tests and uncertainty requirements for instruments whose functions include measuring, recording, and possibly monitoring power quality parameters in power supply systems, and whose measuring methods (class A or class S) are defined in IEC 61000-4-30.

This document applies to power quality instruments complying with IEC 62586-1.

This document can also be referred to by other product standards (e.g. digital fault recorders, revenue meters, MV or HV protection relays) specifying devices embedding class A or class S power quality functions according to IEC 61000-4-30.

These requirements are applicable in single-, dual- (split phase) and 3-phase AC power supply systems at 50 Hz or 60 Hz.

It is not the intent of this document to address user interface or topics unrelated to device measurement performance.

The document does not cover post-processing and interpretation of the data, for example with dedicated software.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61000-2-4, *Electromagnetic compatibility (EMC) – Part 2-4: Environment – Compatibility levels in industrial plants for low-frequency conducted disturbances*

IEC 61000-4-7, *Electromagnetic compatibility (EMC) – Part 4-7: Testing and measurement techniques – General guide on harmonics and interharmonics measurements and instrumentation, for power supply systems and equipment connected thereto*

IEC 61000-4-15, *Electromagnetic compatibility (EMC) – Part 4-15: Testing and measurement techniques – Flickermeter – Functional and design specifications*

IEC 61000-4-30:2015, *Electromagnetic compatibility (EMC) – Part 4-30: Testing and measurement techniques – Power quality measurement methods*

IEC 62586-1:2013, *Power quality measurement in power supply systems – Part 1: Power quality instruments (PQI)*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*