

© Copyright SEK Svensk Elstandard. Reproduction in any form without permission is prohibited.

Elektrisk utrustning för medicinskt bruk – Del 2-60: Särskilda fordringar på säkerhet och väsentliga prestanda för tandvårdsutrustning

Medical electrical equipment –

Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment

Som svensk standard gäller europastandarden EN IEC 80601-2-60:2020. Den svenska standarden innehåller den officiella engelska språkversionen av EN IEC 80601-2-60:2020.

Nationellt förord

Europastandarden EN IEC 80601-2-60:2020

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 80601-2-60, Second edition, 2019 - Medical electrical equipment - Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment**

utarbetad inom International Electrotechnical Commission, IEC.

Standarden ska användas tillsammans med SS-EN 60601-1, utgåva 2.1, 2013.

Tidigare fastställd svensk standard SS-EN 80601-2-60, utgåva 1, 2015, gäller ej fr o m 2023-04-03.

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284
164 29 Kista
Tel 08-444 14 00
www.elstandard.se

EUROPEAN STANDARD

EN IEC 80601-2-60

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2020

ICS 11.040.01

Supersedes EN 80601-2-60:2015 and all of its amendments and corrigenda (if any)

English Version

**Medical electrical equipment - Part 2-60: Particular requirements
for the basic safety and essential performance of dental
equipment
(IEC 80601-2-60:2019)**

Appareils électromédicaux - Partie 2-60: Exigences
particulères pour la sécurité de base et les performances
essentielles des équipements dentaires
(IEC 80601-2-60:2019)

Medizinische elektrische Geräte - Teil 2-60: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Dental-Geräten
(IEC 80601-2-60:2019)

This European Standard was approved by CENELEC on 2019-08-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2020 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC Members.

Ref. No. EN IEC 80601-2-60:2020 E

SEK Svensk Elstandard

SS-EN IEC 80601-2-60, utg 2:2020

European foreword

The text of document 62D/1683/FDIS, future edition 2 of IEC 80601-2-60, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-60:2020.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-10-03
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2023-04-03

This document supersedes EN 80601-2-60:2015 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association.

Endorsement notice

The text of the International Standard IEC 80601-2-60:2019 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

| | | |
|------------------|------|--|
| IEC 60038 | NOTE | Harmonized as EN 60038 |
| IEC 60601-1-10 | NOTE | Harmonized as EN 60601-1-10 |
| IEC 61810-7:2006 | NOTE | Harmonized as EN 61810-7:2006 (not modified) |
| ISO 7494-2:2015 | NOTE | Harmonized as EN ISO 7494-2:2015 (not modified) |
| ISO 13732-1:2006 | NOTE | Harmonized as EN ISO 13732-1:2008 (not modified) |
| ISO 17664:2017 | NOTE | Harmonized as EN ISO 17664:2017 (not modified) |
| ISO 18397:2016 | NOTE | Harmonized as EN ISO 18397:2016 (not modified) |
| ISO 21530:2004 | NOTE | Harmonized as EN ISO 21530:2004 (not modified) |

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Clause 2 of EN 60601-1:2006 applies, except as follows:

| <u>Publication</u> | <u>Year</u> | <u>Title</u> | <u>EN/HD</u> | <u>Year</u> |
|--------------------------------|-------------|--|--|--------------------------------------|
| <i>Replace</i> IEC 60825-1 | 2014 | Safety of laser products - Part 1: Equipment classification and requirements | 1:EN 60825-1 +EN 60825-1:2014/AC:2017-06 +A11 | 2014 2020 |
| <i>Addition</i> IEC 60601-1 | 2005 | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance | 1:EN 60601-1 +A12 +EN 60601-1:2006/corrigendum Mar. 2010 +AC +A11 | 2006 2014 2010 2014 2011 |
| IEC 60601-2-2 | 2017 | Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories | 2:EN IEC 60601-2-2 | 2018 |
| IEC 60601-2-22 | 2007 | Medical electrical equipment - Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment | 22:EN 60601-2-22 | 2013 |
| IEC 60601-2-57 | 2011 | Medical electrical equipment - Part 2-57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use | 57:EN 60601-2-57 | 2011 |
| IEC 60664-1 | 2007 | Insulation coordination for equipment within low-voltage systems - Part 1: Principles, requirements and tests | EN 60664-1 | 2007 |

EN IEC 80601-2-60:2020 (E)

| <u>Publication</u> | <u>Year</u> | <u>Title</u> | <u>EN/HD</u> | <u>Year</u> |
|--------------------|-------------|---|---|-------------|
| IEC 60664-4 | 2005 | Insulation coordination for equipment within low-voltage systems - Part 4: Consideration of high-frequency voltage stress | EN 60664-4 | 2006 |
| | | | +EN 60664-2006 4:2006/corrigendum Oct. 2006 | 2006 |
| IEC 61180 | 2016 | High-voltage test techniques for low-voltage equipment - Definitions, test and procedure requirements, test equipment | EN 61180 | 2016 |
| IEC 61810-1 | 2015 | Electromechanical elementary relays - Part 1: General and safety requirements | EN 61810-1 | 2015 |
| ISO 1942 | 2009 | Dentistry -- Vocabulary | EN ISO 1942 | 2010 |
| ISO 14457 | 2017 | | EN ISO 14457 | 2017 |

CONTENTS

| | |
|--|----|
| FOREWORD..... | 3 |
| 201.1 Scope, object and related standards | 6 |
| 201.2 Normative references..... | 8 |
| 201.3 Terms and definitions..... | 8 |
| 201.4 General requirements | 10 |
| 201.5 General requirements for testing of ME EQUIPMENT | 10 |
| 201.6 Classification of ME EQUIPMENT and ME SYSTEMS..... | 10 |
| 201.7 ME EQUIPMENT identification, marking and documents | 10 |
| 201.8 Protection against electrical HAZARDS from ME EQUIPMENT | 11 |
| 201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS | 16 |
| 201.10 Protection against unwanted and excessive radiation HAZARDS | 19 |
| 201.11 Protection against excessive temperatures and other HAZARDS | 19 |
| 201.12 Accuracy of controls and instruments and protection against hazardous outputs | 23 |
| 201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT | 23 |
| 201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)..... | 24 |
| 201.15 Construction of ME EQUIPMENT | 24 |
| 201.16 ME SYSTEMS | 25 |
| 201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS | 25 |
| 201.101 Cordless HAND-HELD and foot-operated control devices | 25 |
| Annexes | 26 |
| Annex AA (informative) Particular guidance and rationale | 27 |
| Bibliography..... | 39 |
| Index of defined terms used in this document | 40 |
| | |
| Figure AA.1 – Example of APPLIED PARTS for DENTAL EQUIPMENT | 28 |
| Figure AA.2 – Calculation of LEAKAGE CURRENT | 29 |
| Figure AA.3 – Insulation problem of commutator DENTAL ELECTRICAL MOTOR..... | 31 |
| Figure AA.4 – Loading fan construction..... | 37 |
| Figure AA.5 – Load diagram with loading fan | 37 |
| | |
| Table 201.101 – Test voltages for solid insulation for SECONDARY CIRCUITS according to 201.8.9.1.12 | 12 |
| Table 201.102 – Determination of TENSILE SAFETY FACTOR..... | 18 |
| Table 201.103 – Mass distribution..... | 19 |
| Table 201.104 – Allowable maximum temperatures for the OPERATOR SIDE of DENTAL HANDPIECES | 20 |
| Table AA.1 – RATED impulse voltage for equipment energized directly from the low- voltage mains | 32 |

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-60: Particular requirements for the basic safety
and essential performance of dental equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-60 has been prepared by a Joint Working Group of subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and subcommittee 6: Dental equipment, of ISO technical committee 106: Dentistry.

This second edition cancels and replaces the first edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012.

The text of this International Standard is based on the following documents:

| | |
|---------------|------------------|
| FDIS | Report on voting |
| 62D/1683/FDIS | 62D/1691/RVD |

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

This publication is published as a double logo standard.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type*;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE OF DENTAL UNITS, DENTAL PATIENT CHAIRS, DENTAL HANDPIECES AND DENTAL OPERATING LIGHTS, hereafter referred to as DENTAL EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for DENTAL EQUIPMENT (as defined in 201.3.202.)

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60825-1:2014, *Safety of laser products – Part 1: Equipment classification and requirements*

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-2:2017, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-22:2007, *Medical electrical equipment – Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*
IEC 60601-2-22:2007/AMD1:2012

IEC 60601-2-57:2011, *Medical electrical equipment – Part 2-57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use*

IEC 60664-1:2007, *Insulation coordination for equipment within low-voltage systems – Part 1: Principles, requirements and tests*

IEC 60664-4:2005, *Insulation coordination for equipment within low-voltage systems – Part 4: Consideration of high-frequency voltage stress*

IEC 61180:2016, *High-voltage test techniques for low-voltage equipment – Definitions, test and procedure requirements, test equipment*

IEC 61810-1:2015, *Electromechanical elementary relays – Part 1: General and safety requirements*

ISO 1942:2009, *Dentistry – Vocabulary*

ISO 14457: 2017, *Dentistry – Handpieces and motors*