
Medical electrical equipment —
Part 2-72:
Particular requirements for basic
safety and essential performance
of home healthcare environment
ventilators for ventilator-dependent
patients

Appareils électromédicaux —

*Partie 2-72: Exigences particulières pour la sécurité de base
et les performances essentielles des ventilateurs utilisés dans
l'environnement des soins à domicile pour les patients ventilo-
dépendants*





COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Foreword	v
Introduction.....	vii
201.1 Scope, object, and related standards.....	1
201.2 Normative references	3
201.3 Terms and definitions	5
201.4 General requirements	21
201.5 General requirements for testing of <i>ME equipment</i>	25
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	26
201.7 <i>ME equipment</i> identification, marking, and documents	26
201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>	35
201.9 Protection against mechanical hazards of <i>ME equipment</i> and <i>ME systems</i>	35
201.10 Protection against unwanted and excessive radiation <i>hazards</i>	37
201.11 Protection against excessive temperatures and other <i>hazards</i>	38
201.12 Accuracy of controls and instruments and protection against hazardous outputs	40
201.13 <i>Hazardous situations</i> and fault conditions.....	59
201.14 Programmable electrical medical systems (PEMS)	61
201.15 Construction of <i>ME equipment</i>	61
201.16 <i>ME systems</i>	62
201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i>	63
201.101 Gas connections	63
201.102 Requirements for the <i>VBS</i> and <i>accessories</i>	66
201.103 Spontaneous breathing during loss of power supply.....	67
201.104 Indication of duration of operation	68
201.105 Functional connection	68
201.106 Display loops.....	69
201.107 <i>Ventilator</i> security	69
201.108 Oxygen inlet port.....	70
202 Electromagnetic disturbances — Requirements and tests.....	70
206 Usability.....	71
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	73
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.....	75

Annex C (informative) Guide to marking and labelling requirements for ME equipment and ME systems.....	78
Annex D (informative) Symbols on marking.....	85
Annex AA (informative) Particular guidance and rationale.....	87
Annex BB (informative) Data interface requirements	113
Annex CC (informative) Reference to the IMDRF essential principles and labelling guidances.....	121
Annex DD (informative) Reference to the essential principles.....	126
Annex EE (informative) Reference to the general safety and performance requirements.....	130
Bibliography	134

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see <https://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care* and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC 62D, *Electric equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 80601-2-72:2015), which has been technically revised.

The main changes are as follows:

- added requirements for display during calibration of gas monitors;
- clarified *maximum limited pressure* requirements;
- clarified high *airway pressure alarm condition* requirements;
- added requirements for *ventilator system recovery*;
- added requirements for response to an increase in set oxygen (O₂) concentration; and
- harmonization with ISO 20417, where appropriate.

A list of all parts in the ISO 80601 series and the IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

Introduction

This document specifies requirements for *lung ventilators* that are intended for use in the *home healthcare environment* for *patients* who are dependent on *ventilation* for their life support. These *ventilators* are frequently used in locations where the *supply mains* driving the *ventilator* is not reliable. These *ventilators* are often supervised by non-healthcare personnel (*lay operators*) with varying levels of training. *Lung ventilators* conforming with this standard can be used elsewhere (i.e. in healthcare facilities).

In referring to the structure of this document,

- “clause” means one of the 5 numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes 201.7, 201.8, etc.); and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the *marking* and labelling requirements in this document.

Annex D contains a summary of the *symbols* referenced in this document.

Requirements in this document have been decomposed so that each requirement is uniquely delineated. This is done to support automated requirements tracking.

Medical electrical equipment —

Part 2-72:

Particular requirements for basic safety and essential performance of home healthcare environment ventilators for ventilator-dependent patients

201.1 Scope, object, and related standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 1 applies, except as follows:

201.1.1 Scope

Replacement:

NOTE 1 There is guidance or rationale for this subclause contained in Clause AA.2.

This document applies to the *basic safety* and *essential performance* of a *ventilator* in combination with its *accessories*, hereafter referred to as *ME equipment*:

— intended for use in the *home healthcare environment*;

NOTE 2 In the *home healthcare environment*, the *supply mains* driving the *ventilator* is often not reliable.

NOTE 3 Such *ventilators* can also be used in non-critical care applications of *professional healthcare facilities*.

— intended for use by a *lay operator*; and

— intended for those *patients* who need differing levels of support from *artificial ventilation* including for *ventilator-dependent patients*.

A *ventilator* is not considered to use a *physiologic closed-loop control system* unless it uses a physiological *patient* variable to adjust the *ventilation* therapy settings.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to a *ventilator breathing system* or to a *ventilator* where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *ventilator*.

EXAMPLE Breathing tubes, *connectors*, water traps, expiratory valve, *humidifier*, *breathing system filter*, external electrical power source, and *distributed alarm system*.

NOTE 4 If a clause or subclause is specifically intended to be applicable to *ME equipment* only or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except for the requirements specified in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 7.2.13 and 8.4.1.

NOTE 5 Additional information can be found in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 4.2.

This document does not specify the requirements for:

- *ventilators* or *accessories* intended for critical care applications, which are given in ISO 80601-2-12;
- *ventilators* or *accessories* intended for anaesthetic applications, which are given in ISO 80601-2-13;
- *ventilators* or *accessories* intended for emergency and transport which are given in ISO 80601-2-84;
- *ventilators* or *accessories* intended for homecare ventilatory support equipment (intended only to augment the *ventilation* of spontaneously breathing *patients*), which are given in ISO 80601-2-79 and ISO 80601-2-80;
- obstructive sleep apnoea therapy *ME equipment*, which are given in ISO 80601-2-70;
- high-frequency *ventilators*, which are given in ISO 80601-2-87.
- respiratory high-flow therapy equipment, which are given in ISO 80601-2-90;

NOTE 6 An ISO 80601-2-72 *ventilator* can incorporate high-flow therapy operational mode, but such a mode is only for spontaneously breathing *patients*.

- user-powered resuscitators, which are given in ISO 10651-4;
- gas-powered emergency resuscitators, which are given in ISO 10651-5;
- oxygen therapy constant flow *ME equipment*; and
- cuirass and “iron-lung” *ventilators*.