

© Copyright SEK Svensk Elstandard. Reproduction in any form without permission is prohibited.

REDLINE VERSION

Elektrisk utrustning för medicinskt bruk – Säkerhet och väsentliga prestanda – Del 2-43: Särskilda fordringar på röntgenutrustning för interventionella procedurer

Medical electrical equipment –

Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures

En så kallad ”Redline version” (RLV) innehåller både den fastställda IEC-standarden och en ändringsmarkerad standard. Alla tillägg och borttagningar sedan den tidigare utgåvan är markerade med färg. Med en RLV sparar du mycket tid när du ska identifiera och bedöma aktuella ändringar i standarden. SEK Svensk Elstandard kan bara ge ut RLV i de fall den finns tillgänglig från IEC.



IEC 60601-2-43

Edition 3.0 2022-12
REDLINE VERSION

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-43: Particular requirements for the basic safety and essential
performance of X-ray equipment for interventional procedures**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.50; 37.040.25

ISBN 978-2-8322-6213-9

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	8
201.2 Normative references.....	11
201.3 Terms and definitions.....	12
201.4 General requirements	13
201.5 General requirements for testing ME EQUIPMENT	15
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	15
201.7 ME EQUIPMENT identification, marking and documents	15
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	20
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	20
201.10 Protection against unwanted and excessive radiation HAZARDS	22
201.11 Protection against excessive temperatures and other HAZARDS	22
201.12 Accuracy of controls and instruments and protection against hazardous outputs	24
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	27
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	27
201.15 Construction of ME EQUIPMENT.....	28
201.16 ME SYSTEMS	28
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	28
202 Electromagnetic disturbances – Requirements and tests	29
203 RADIATION PROTECTION in diagnostic X-RAY EQUIPMENT	29
Annexes	42
Annex AA (informative) Particular guidance and rationale.....	43
Annex BB (normative) Distribution maps of STRAY RADIATION.....	55
Annex CC (informative) Mapping between this Edition 2 of IEC 60601-2-43 and Edition 1
Bibliography.....	60
Index of defined terms	63
Figure BB.1 – Example of isokerma map at 100 cm height in lateral configuration	56
Figure BB.2 – Example of isokerma map at 100 cm height in vertical configuration.....	57
Table 201.101 – Additional list of potential ESSENTIAL PERFORMANCE to be considered by the MANUFACTURER in the RISK MANAGEMENT analysis	13
Table 201.102 – Other subclauses requiring statements in ACCOMPANYING DOCUMENTS	20
Table AA.1 – Examples of prolonged RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES for which deterministic effects of IRRADIATION are possible	43
Table AA.2 – Examples of RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES for which deterministic effects are unlikely	44
Table AA.3 – Examples of isodose boundaries and colour codes for SKIN DOSE MAP and AIR KERMA map.....	53

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition IEC 60601-2-43:2010+AMD1:2017+AMD2:2019 CSV. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

IEC 60601-2-43 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice. It is an International Standard.

This third edition cancels and replaces the second edition published in 2010, Amendment 1:2017 and Amendment 2:2019. This edition constitutes a technical revision.

This edition includes editorial and technical changes to reflect the changes in IEC 60601-1:2005/AMD2:2020 and IEC 60601-2-54:2022. It also contains corrections and technical improvements. Significant technical changes with respect to the previous edition are as follows:

- a) a new specific term DOSIMETER is introduced to replace the general term DOSEMETER as in IEC 60601-2-54:2022;
- b) several terms and definitions that are moved from IEC TR 60788:2004 to 201.3 of IEC 60601-2-54:2022 are also referenced from IEC 60601-2-54:2022.
- c) the collateral standards IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 are applicable if MANUFACTURER so declares;
- d) the former subclause 201.11.101 “Protection against excessive temperature of X-RAY TUBE ASSEMBLIES” is removed since covered by IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020 and IEC 60601-2-28:2017, and the former subclause 201.11.102 is renumbered as 201.11.101, as in IEC 60601-2-54:2022;
- e) to adopt changes in subclause 7.8.1 “Colours of indicator lights” in IEC 60601-1:2005/AMD2:2020, clarification of requirements is provided in 201.7.8.1 to avoid conflicts with requirements of indicator lights stipulated for X-RAY EQUIPMENT, as in IEC 60601-2-54:2022;
- f) explanation of the term ESSENTIAL PERFORMANCE is provided in Annex AA to emphasize the performance of the clinical function under NORMAL CONDITIONS and SINGLE FAULT CONDITIONS.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62B/1297/FDIS	62B/1309/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED in Clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, in this document or as noted: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 and IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

<p>IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.</p>
--

INTRODUCTION

The purpose of this new edition is to introduce changes to reference the Amendment 2 (2020) to IEC 60601-1:2005 and some minor technical clarifications.

X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES ~~may~~ can subject PATIENTS and OPERATORS to higher levels of RADIATION than those which normally prevail during diagnostic X-ray imaging procedures. One consequence for the PATIENT ~~may~~ can be the occurrence of deterministic injury when RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES involve the delivery of substantial amounts of RADIATION to localized areas. Another consequence can be an increased RISK of stochastic effects, such as cancer. These health concerns apply also to the OPERATOR. In addition, for this particular type of equipment, there is a need for availability of critical functions with minimal periods of loss.

RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES of the type envisaged are well established in clinical fields such as:

- invasive cardiology;
- interventional RADIOLOGY;
- interventional neuroradiology.

These RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES also include many newly developing and emerging applications in a wide range of medical and surgical specialties.

NOTE Attention is drawn to the existence of legislation in some countries concerning RADIOLOGICAL PROTECTION, which ~~may~~ sometimes do not align with the provisions of this document.

~~INTRODUCTION to Amendment 1~~

~~The purpose of this first amendment to IEC 60601-2-43:2010 is to introduce changes as follows:~~

- ~~— refer to IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 and its applicable collateral standards;~~
- ~~— refer to IEC 60601-2-54:2009 and IEC 60601-2-54:2009/AMD1:2015 and consequent subclause adaptations;~~
- ~~— include a requirement to have a maximum time of 10 min to recover all functions after a recoverable failure in 201.4.101;~~
- ~~— include several aspects from IEC 61910-1:2014 and remove the reference to IEC PAS 61910-1:2007 in 201.4.102;~~
- ~~— include an alternative way of testing in 201.11.6.5.103;~~
- ~~— include a clarification for tableside controls in 201.12.4.106.~~

~~In addition, a number of technical errors have been corrected.~~

~~INTRODUCTION to Amendment 2~~

~~The purpose of this second amendment to IEC 60601-2-43:2010 is to introduce changes as follows:~~

- ~~— scope clarification with regards to MOBILE X-ray equipment and applicability of IEC 60601-2-54 subclauses;~~
- ~~— reference to IEC 60601-2-54:2009/AMD2:2018 for common subclauses;~~

- ~~— alignment of 201.7.9.1 with IEC 60601-2-54:2009/AMD2:2018 — 201.7.9.1 is no longer modified;~~
- ~~— inclusion of adapted requirements or recommendations from IEC 60601-2-54:2009/AMD2:2018 for~~
 - ~~• management of radiology image storage in 203.6.1.101,~~
 - ~~• display of last image hold (LIH RADIOGRAM) in 203.6.7.101, and~~
 - ~~• graphical indication of the boundaries of the X-RAY FIELD in 203.8.102.2;~~
- ~~— inclusion of a recommendation for protection of gantry enclosures in 201.11.6.5.103;~~
- ~~— inclusion of a requirement for X-RADIATION pulse repetition frequency during radiology in 203.6.3.103;~~
- ~~— inclusion of a recommendation for a DOSE MAP in 203.6.4.5 with additional definitions in 201.3;~~
- ~~— inclusion of a requirement for display unit of dose area product in 203.6.4.5;~~
- ~~— addition of a number of technical clarifications.~~

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures

201.1 Scope, object and related standards

Clause 1 of ~~the general standard~~¹⁾ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 * Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of both FIXED and MOBILE X-RAY EQUIPMENT declared by the MANUFACTURER to be suitable for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, hereafter referred to as INTERVENTIONAL X-RAY EQUIPMENT. Its scope excludes, in particular:

- equipment for RADIOTHERAPY;
- equipment for COMPUTED TOMOGRAPHY;
- ACCESSORIES intended to be introduced into the PATIENT;
- mammographic X-RAY EQUIPMENT;
- dental X-RAY EQUIPMENT.

NOTE 1 Examples of RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, for which the use of INTERVENTIONAL X-RAY EQUIPMENT complying with this document is recommended, are given in Annex AA.

NOTE 2 Specific requirements for magnetic navigation devices, and for the use of INTERVENTIONAL X-RAY EQUIPMENT in an operating room environment were not considered in this document; therefore, no specific requirements have been developed for these devices or uses. In any case, such devices or uses remain under the general clause requirements.

NOTE 3 INTERVENTIONAL X-RAY EQUIPMENT, when used for cone-beam CT mode, is covered by this document and not by IEC 60601-2-44 [1]². No additional requirements for operation in cone-beam CT mode were identified for this document (see also Note 5 in 203.6.4.5).

INTERVENTIONAL X-RAY EQUIPMENT declared by the MANUFACTURER to be suitable for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, which does not include a PATIENT SUPPORT as part of the system, is exempt from the PATIENT SUPPORT provisions of this document.

If a clause or subclause is specifically intended to be applicable to INTERVENTIONAL X-RAY EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to INTERVENTIONAL X-RAY EQUIPMENT and to ME SYSTEMS, as relevant.

~~NOTE 4 See also 4.2 of the general standard.~~

¹⁾ ~~The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.~~

² Numbers in square brackets refer to the Bibliography.

~~The subclauses of this standard supersede IEC 60601-2-54 subclauses.~~ IEC 60601-2-54 applies only with regards to the cited subclauses; non-cited subclauses of IEC 60601-2-54 do not apply.

201.1.2 Object

Replacement:

The object of this document is:

- to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for the design and manufacture of X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, as defined in 201.3.205.
- to specify information which ~~is to~~ shall be provided with such INTERVENTIONAL X-RAY EQUIPMENT for the assistance of the RESPONSIBLE ORGANIZATION and OPERATOR in managing the RADIATION RISK and equipment failure RISK arising from these RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES which could affect PATIENTS or staff.

201.1.3 Collateral standards

Addition:

~~This document refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.~~

~~IEC 60601-1-2 and IEC 60601-1-3 apply as modified in Clause 202 and Clause 203 respectively.~~

~~IEC 60601-1-8³⁾, IEC 60601-1-9⁴⁾, IEC 60601-1-10⁵⁾, IEC 60601-1-11⁶⁾ and IEC 60601-1-12⁷⁾ do not apply.~~

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, as modified in 201.2.

IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-3:2008, IEC 60601-1-3:2008/AMD1:2013 and IEC 60601-1-3:2008/AMD2:2021 apply as modified in Clause 202 and Clause 203 respectively.

IEC 60601-1-8 [2], IEC 60601-1-9 [3], IEC 60601-1-10 [4] do not apply.

NOTE OPERATORS of INTERVENTIONAL X-RAY EQUIPMENT are used to audible signals as specified in this document rather than the concepts of IEC 60601-1-8.

~~3) IEC 60601-1-8, Medical electrical equipment — Part 1-8: General requirements for basic safety and essential performance — Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems~~

~~4) IEC 60601-1-9, Medical electrical equipment — Part 1-9: General requirements for basic safety and essential performance — Collateral Standard: Requirements for environmentally conscious design~~

~~5) IEC 60601-1-10, Medical electrical equipment — Part 1-10: General requirements for basic safety and essential performance — Collateral Standard: Requirements for the development of physiologic closed-loop controllers~~

~~6) IEC 60601-1-11, Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment~~

~~7) IEC 60601-1-12, Medical electrical equipment — Part 1-12: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment~~

IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020 [5] apply only if the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in a HOME HEALTHCARE ENVIRONMENT, and otherwise do not apply.

IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 [6] apply only if the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in an EMERGENCY MEDICAL SERVICES ENVIRONMENT, and otherwise do not apply.

All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

~~For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.~~

The numbering of clauses and subclauses of this particular standard corresponds to that of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 addresses the content of Clause 1 of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x.101" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the ~~general~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard.

"*Amendment*" means that the clause or subclause of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However, due to the fact that definitions in ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.439154, additional definitions are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

~~The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.~~

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the ~~general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the ~~general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

NOTE Informative references are listed in the Bibliography.

Amendment:

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*
IEC 60529:1989/AMD1:1999
IEC 60529:1989/AMD2:2013

~~IEC 60601-1-2:2014 *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic disturbances – Requirements and tests*~~

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment*
IEC 60601-1-3:2008/AMD1:2013
IEC 60601-1-3:2008/AMD2:2021

~~Delete the reference to IEC 60601-1-8 and its amendments.~~

Addition:

IEC 60580:20002019, *Medical electrical equipment – Dose area product meters*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012
IEC 60601-1:2005/AMD2:2020

IEC 60601-2-54:20092022, *Medical electrical equipment – Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy*
~~IEC 60601-2-54:2009/AMD1:2015~~
~~IEC 60601-2-54:2009/AMD2:2018~~

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 61910-1:2014, *Medical electrical equipment – Radiation dose documentation – Part 1: Radiation dose structured reports for radiography and radioscopy*

~~IEC 62220-1:2003, *Medical electrical equipment – Characteristics of digital X-ray imaging devices – Part 1: Determination of the detective quantum efficiency*~~

IEC 62220-1-1:2015, *Medical electrical equipment – Characteristics of digital X-ray imaging devices – Part 1-1: Determination of the detective quantum efficiency – Detectors used in radiographic imaging*

ISO 14971, *Medical devices – Application of risk management to medical devices*

© Copyright SEK Svensk Elstandard. Reproduction in any form without permission is prohibited.

Elektrisk utrustning för medicinskt bruk – Säkerhet och väsentliga prestanda – Del 2-43: Särskilda fordringar på röntgenutrustning för interventionella procedurer

Medical electrical equipment –

*Part 2-43: Particular requirements for the basic safety and essential performance
of X-ray equipment for interventional procedures*

Som svensk standard gäller europastandarden EN IEC 60601-2-43:2023. Den svenska standarden innehåller den officiella engelska språkversionen av EN IEC 60601-2-43:2023.

Nationellt förord

Europastandarden EN IEC 60601-2-43:2023

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 60601-2-43, Third edition, 2022 - Medical electrical equipment –
Part 2-43: Particular requirements for the basic
safety and essential performance of X-ray equipment
for interventional procedures**

utarbetad inom International Electrotechnical Commission, IEC.

ICS 11.040.50; 37.040.25

Denna standard är fastställd av SEK Svensk Elstandard,
som också kan lämna upplysningar om **sakinnehållet** i standarden.
Postadress: Box 1284, 164 29 KISTA
Telefon: 08 - 444 14 00.
E-post: sek@elstandard.se. Internet: www.elstandard.se

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284
164 29 Kista
Tel 08-444 14 00
www.elstandard.se

EUROPEAN STANDARD

EN IEC 60601-2-43

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2023

ICS 11.040.50; 37.040.25

Supersedes EN 60601-2-43:2010;
EN 60601-2-43:2010/AC:2014;
EN 60601-2-43:2010/A1:2018;
EN 60601-2-43:2010/A2:2020

English Version

**Medical electrical equipment - Part 2-43: Particular requirements
for the basic safety and essential performance of X-ray
equipment for interventional procedures
(IEC 60601-2-43:2022)**

Appareils électromédicaux - Partie 2-43: Exigences
particulières pour la sécurité de base et les performances
essentielle des appareils à rayonnement X lors
d'interventions
(IEC 60601-2-43:2022)

Medizinische elektrische Geräte - Teil 2-43: Besondere
Festlegungen für die Sicherheit und wesentlichen
Leistungsmerkmale von Röntgeneinrichtungen für
interventionelle Verfahren
(IEC 60601-2-43:2022)

This European Standard was approved by CENELEC on 2023-01-09. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2023 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC Members.

Ref. No. EN IEC 60601-2-43:2023 E

SEK Svensk Elstandard

SS-EN IEC 60601-2-43, utg 3:2023

European foreword

The text of document 62B/1297/FDIS, future edition 3 of IEC 60601-2-43, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-43:2023.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2023-10-09
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2026-01-09

This document supersedes EN 60601-2-43:2010 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 60601-2-43:2022 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60601-2-44 NOTE Harmonized as EN 60601-2-44

IEC 60601-1-8 NOTE Harmonized as EN 60601-1-8

IEC 60601-1-9 NOTE Harmonized as EN 60601-1-9

IEC 60601-1-10 NOTE Harmonized as EN 60601-1-10

IEC 60601-1-11:2015 NOTE Harmonized as EN 60601-1-11:2015 (not modified)

IEC 60601-1-12:2014 NOTE Harmonized as EN 60601-1-12:2015 (not modified)

IEC 60601-2-28:2017 NOTE Harmonized as EN IEC 60601-2-28:2019 (not modified)

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Clause 2 of EN 60601-1:2006¹, EN 60601-1:2006/A1:2013, and EN 60601-1:2005/A2:2021 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
--------------------	-------------	--------------	--------------	-------------

Replace the references to IEC 60529:1989 and IEC 60601-1-3 with the following new references:

IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	EN 60529	1991
-	-		+ corrigendum	May 1993
+ A1	1999		+ A1	2000
+ A2	2013		+ A2	2013
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3	2008
			+ corrigendum	Mar. 2010
+ A1	2013		+ A1	2013
			+ AC	2014
			+ A11	2016
+ A2	2021		+ A2	2021

Delete the reference to IEC 60601-1-8 and its amendments.

¹ As impacted by EN 60601-1:2006/corrigendum Mar. 2010, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A1:2013/AC:2014 and EN 60601-1:2006/AC:2022-12.

EN IEC 60601-2-43:2023 (E)

Add the following references:

IEC 60580	2019	Medical electrical equipment - Dose area product meters	EN IEC 60580	2020
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-		+ corrigendum Mar.	2010
+ A1	2012		+ A1	2013
-	-		+ A12	2014
+ A2	2020		+ A2	2021
IEC 60601-2-54	2022	Medical electrical equipment - Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy	-	-
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 61910-1	2014	Medical electrical equipment - Radiation dose documentation - Part 1: Radiation dose structured reports for radiography and radioscopy	EN 61910-1	2014
IEC 62220-1-1	2015	Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1-1: Determination of the detective quantum efficiency - Detectors used in radiographic imaging	EN 62220-1-1	2015
ISO 14971	-	Medical devices - Application of risk management to medical devices	-	-

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Medical electrical equipment –

Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures

Appareils électromédicaux –

Partie 2-43: Exigences particulières pour la sécurité de base et les performances essentielles des appareils à rayonnement X lors d'interventions

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50; 37.040.25

ISBN 978-2-8322-6087-6

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	7
201.2 Normative references.....	9
201.3 Terms and definitions.....	10
201.4 General requirements	11
201.5 General requirements for testing ME EQUIPMENT	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	13
201.7 ME EQUIPMENT identification, marking and documents	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	17
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	18
201.10 Protection against unwanted and excessive radiation HAZARDS	19
201.11 Protection against excessive temperatures and other HAZARDS	19
201.12 Accuracy of controls and instruments and protection against hazardous outputs	21
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	24
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	24
201.15 Construction of ME EQUIPMENT.....	25
201.16 ME SYSTEMS	25
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	25
202 Electromagnetic disturbances – Requirements and tests.....	26
203 RADIATION PROTECTION in diagnostic X-RAY EQUIPMENT	26
Annexes	39
Annex AA (informative) Particular guidance and rationale.....	40
Annex BB (normative) Distribution maps of STRAY RADIATION.....	52
Bibliography.....	55
Index of defined terms	58
Figure BB.1 – Example of isokerma map at 100 cm height in lateral configuration	53
Figure BB.2 – Example of isokerma map at 100 cm height in vertical configuration.....	54
Table 201.101 – Additional list of potential ESSENTIAL PERFORMANCE to be considered by the MANUFACTURER in the RISK MANAGEMENT analysis	11
Table 201.102 – Other subclauses requiring statements in ACCOMPANYING DOCUMENTS	17
Table AA.1 – Examples of prolonged RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES for which deterministic effects of IRRADIATION are possible	40
Table AA.2 – Examples of RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES for which deterministic effects are unlikely	41
Table AA.3 – Examples of isodose boundaries and colour codes for SKIN DOSE MAP and AIR KERMA map.....	50

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-43 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice. It is an International Standard.

This third edition cancels and replaces the second edition published in 2010, Amendment 1:2017 and Amendment 2:2019. This edition constitutes a technical revision.

This edition includes editorial and technical changes to reflect the changes in IEC 60601-1:2005/AMD2:2020 and IEC 60601-2-54:2022. It also contains corrections and technical improvements. Significant technical changes with respect to the previous edition are as follows:

- a) a new specific term DOSIMETER is introduced to replace the general term DOSEMETER as in IEC 60601-2-54:2022;
- b) several terms and definitions that are moved from IEC TR 60788:2004 to 201.3 of IEC 60601-2-54:2022 are also referenced from IEC 60601-2-54:2022.

- c) the collateral standards IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 are applicable if MANUFACTURER so declares;
- d) the former subclause 201.11.101 "Protection against excessive temperature of X-RAY TUBE ASSEMBLIES" is removed since covered by IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020 and IEC 60601-2-28:2017, and the former subclause 201.11.102 is renumbered as 201.11.101, as in IEC 60601-2-54:2022;
- e) to adopt changes in subclause 7.8.1 "Colours of indicator lights" in IEC 60601-1:2005/AMD2:2020, clarification of requirements is provided in 201.7.8.1 to avoid conflicts with requirements of indicator lights stipulated for X-RAY EQUIPMENT, as in IEC 60601-2-54:2022;
- f) explanation of the term ESSENTIAL PERFORMANCE is provided in Annex AA to emphasize the performance of the clinical function under NORMAL CONDITIONS and SINGLE FAULT CONDITIONS.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62B/1297/FDIS	62B/1309/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED in Clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, in this document or as noted: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 and IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

The purpose of this new edition is to introduce changes to reference the Amendment 2 (2020) to IEC 60601-1:2005 and some minor technical clarifications.

X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES can subject PATIENTS and OPERATORS to higher levels of RADIATION than those which normally prevail during diagnostic X-ray imaging procedures. One consequence for the PATIENT can be the occurrence of deterministic injury when RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES involve the delivery of substantial amounts of RADIATION to localized areas. Another consequence can be an increased RISK of stochastic effects, such as cancer. These health concerns apply also to the OPERATOR. In addition, for this particular type of equipment, there is a need for availability of critical functions with minimal periods of loss.

RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES of the type envisaged are well established in clinical fields such as:

- invasive cardiology;
- interventional RADIOLOGY;
- interventional neuroradiology.

These RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES also include many newly developing and emerging applications in a wide range of medical and surgical specialties.

NOTE Attention is drawn to the existence of legislation in some countries concerning RADIOLOGICAL PROTECTION, which sometimes do not align with the provisions of this document.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures

201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 * Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of both FIXED and MOBILE X-RAY EQUIPMENT declared by the MANUFACTURER to be suitable for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, hereafter referred to as INTERVENTIONAL X-RAY EQUIPMENT. Its scope excludes, in particular:

- equipment for RADIOTHERAPY;
- equipment for COMPUTED TOMOGRAPHY;
- ACCESSORIES intended to be introduced into the PATIENT;
- mammographic X-RAY EQUIPMENT;
- dental X-RAY EQUIPMENT.

NOTE 1 Examples of RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, for which the use of INTERVENTIONAL X-RAY EQUIPMENT complying with this document is recommended, are given in Annex AA.

NOTE 2 Specific requirements for magnetic navigation devices, and for the use of INTERVENTIONAL X-RAY EQUIPMENT in an operating room environment were not considered in this document; therefore, no specific requirements have been developed for these devices or uses. In any case, such devices or uses remain under the general clause requirements.

NOTE 3 INTERVENTIONAL X-RAY EQUIPMENT, when used for cone-beam CT mode, is covered by this document and not by IEC 60601-2-44 [1]¹. No additional requirements for operation in cone-beam CT mode were identified for this document (see also Note 5 in 203.6.4.5).

INTERVENTIONAL X-RAY EQUIPMENT declared by the MANUFACTURER to be suitable for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, which does not include a PATIENT SUPPORT as part of the system, is exempt from the PATIENT SUPPORT provisions of this document.

If a clause or subclause is specifically intended to be applicable to INTERVENTIONAL X-RAY EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to INTERVENTIONAL X-RAY EQUIPMENT and to ME SYSTEMS, as relevant.

IEC 60601-2-54 applies only with regards to the cited subclauses; non-cited subclauses of IEC 60601-2-54 do not apply.

¹ Numbers in square brackets refer to the Bibliography.

201.1.2 Object

Replacement:

The object of this document is:

- to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for the design and manufacture of X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, as defined in 201.3.205.
- to specify information which shall be provided with such INTERVENTIONAL X-RAY EQUIPMENT for the assistance of the RESPONSIBLE ORGANIZATION and OPERATOR in managing the RADIATION RISK and equipment failure RISK arising from these RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES which could affect PATIENTS or staff.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, as modified in 201.2.

IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-3:2008, IEC 60601-1-3:2008/AMD1:2013 and IEC 60601-1-3:2008/AMD2:2021 apply as modified in Clause 202 and Clause 203 respectively.

IEC 60601-1-8 [2], IEC 60601-1-9 [3], IEC 60601-1-10 [4] do not apply.

NOTE OPERATORS of INTERVENTIONAL X-RAY EQUIPMENT are used to audible signals as specified in this document rather than the concepts of IEC 60601-1-8.

IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020 [5] apply only if the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in a HOME HEALTHCARE ENVIRONMENT, and otherwise do not apply.

IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 [6] apply only if the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in an EMERGENCY MEDICAL SERVICES ENVIRONMENT, and otherwise do not apply.

All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

The numbering of clauses and subclauses of this particular standard corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x.101" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard.

"Amendment" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However, due to the fact that definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.154, additional definitions are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

NOTE Informative references are listed in the Bibliography.

Amendment:

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*
IEC 60529:1989/AMD1:1999
IEC 60529:1989/AMD2:2013

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment*

IEC 60601-1-3:2008/AMD1:2013

IEC 60601-1-3:2008/AMD2:2021

Delete the reference to IEC 60601-1-8 and its amendments.

Addition:

IEC 60580:2019, *Medical electrical equipment – Dose area product meters*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC 60601-2-54:2022, *Medical electrical equipment – Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy*

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 61910-1:2014, *Medical electrical equipment – Radiation dose documentation – Part 1: Radiation dose structured reports for radiography and radioscopy*

IEC 62220-1-1:2015, *Medical electrical equipment – Characteristics of digital X-ray imaging devices – Part 1-1: Determination of the detective quantum efficiency – Detectors used in radiographic imaging*

ISO 14971, *Medical devices – Application of risk management to medical devices*