

© Copyright SEK. Reproduction in any form without permission is prohibited.

**Elektrisk utrustning för mätning, styrning och laboratorieändamål –
EMC-fordringar –
Del 3-2: Immunitetsfordringar på system av betydelse för
säkerheten (säkerhetskritiska system) och på utrustning
med säkerhetsfunktion –
Tillämpningar i industrin i specificerad elektromagnetisk miljö**

*Electrical equipment for measurement, control and laboratory use –
EMC requirements –
Part 3-2: Immunity requirements for safety-related systems and for equipment
intended to perform safety related functions (functional safety) –
Industrial applications with specified electromagnetic environment*

Som svensk standard gäller europastandarden EN 61326-3-2:2008. Den svenska standarden innehåller den officiella engelska språkversionen av EN 61326-3-2:2008.

Nationellt förord

Europastandarden EN 61326-3-2:2008

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 61326-3-2, First edition, 2008 - Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 3-2: Immunity requirements for safety-related systems and for equipment intended to perform safety related functions (functional safety) - Industrial applications with specified electromagnetic environment**

utarbetad inom International Electrotechnical Commission, IEC.

Standarden ska användas tillsammans med SS-EN 61326-1.

ICS 25.040.40; 33.100.20

Denna standard är fastställd av SEK Svensk Elstandard, som också kan lämna upplysningar om **sakinnehållet** i standarden.
Postadress: SEK, Box 1284, 164 29 KISTA
Telefon: 08 - 444 14 00. Telefax: 08 - 444 14 30
E-post: sek@elstandard.se. Internet: www.elstandard.se

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a säkerhet, prestanda, dokumentation, utförande och skötsel av elprodukter, elanläggningar och metoder. Genom att utforma sådana standarder blir säkerhetskraven tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1284
164 29 Kista
Tel 08-444 14 00
www.elstandard.se

**Electrical equipment for measurement, control and laboratory use -
EMC requirements -
Part 3-2: Immunity requirements for safety-related systems
and for equipment intended to perform safety-related functions
(functional safety) -
Industrial applications with specified electromagnetic environment
(IEC 61326-3-2:2008)**

Matériel électrique de mesure,
de commande et de laboratoire -
Exigences relatives à la CEM -
Partie 3-2: Exigences d'immunité
pour les systèmes relatifs à la sécurité
et pour les matériels destinés à réaliser des
fonctions relatives à la sécurité
(sécurité fonctionnelle) -
Applications industrielles
dont l'environnement électromagnétique
est spécifié
(CEI 61326-3-2:2008)

Elektrische Mess-, Steuer-,
Regel- und Laborgeräte -
EMV-Anforderungen -
Teil 3-2: Störfestigkeitsanforderungen
für sicherheitsbezogene Systeme und
für Geräte, die für sicherheitsbezogene
Funktionen vorgesehen sind
(Funktionale Sicherheit) -
Industrielle Anwendungen in spezifizierter
elektromagnetischer Umgebung
(IEC 61326-3-2:2008)

This European Standard was approved by CENELEC on 2008-06-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 65A/501/FDIS, future edition 1 of IEC 61326-3-2, prepared by SC 65A, System aspects, of IEC TC 65, Industrial-process measurement, control and automation, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 61326-3-2 on 2008-06-01.

The EN 61326 series supersedes EN 61326:1997 + corrigendum September 1998 + A1:1998 + A2:2001 + A3:2003.

This standard is to be used in conjunction with EN 61326-1.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 2009-03-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2011-06-01

Annex ZA has been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 61326-3-2:2008 was approved by CENELEC as a European Standard without any modification.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60050-161	- ¹⁾	International Electrotechnical Vocabulary (IEV) - Chapter 161: Electromagnetic compatibility	-	-
IEC 61000-4-2 A1 A2	1995 1998 2000	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2 A1 A2	1995 1998 2001
IEC 61000-4-3	2006	Electromagnetic compatibility (EMC) - Part 4-3: Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test	EN 61000-4-3	2006
IEC 61000-4-4	2004	Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques - Electrical fast transient/burst immunity test	EN 61000-4-4	2004
IEC 61000-4-5	2005	Electromagnetic compatibility (EMC) - Part 4-5: Testing and measurement techniques - Surge immunity test	EN 61000-4-5	2006
IEC 61000-4-6 + A1 + A2	2003 2004 2006	Electromagnetic compatibility (EMC) - Part 4-6: Testing and measurement techniques - Immunity to conducted disturbances, induced by radio-frequency fields	EN 61000-4-6 + corr. August	2007 2007
IEC 61000-4-8 A1	1993 2000	Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test	EN 61000-4-8 A1	1993 2001
IEC 61000-4-11	2004	Electromagnetic compatibility (EMC) - Part 4-11: Testing and measurement techniques - Voltage dips, short interruptions and voltage variations immunity tests	EN 61000-4-11	2004
IEC 61000-4-29	2000	Electromagnetic compatibility (EMC) - Part 4-29: Testing and measurement techniques - Voltage dips, short interruptions and voltage variations on d.c. input power port immunity tests	EN 61000-4-29	2000
IEC 61000-6-2	2005	Electromagnetic compatibility (EMC) - Part 6-2: Generic standards - Immunity for industrial environments	EN 61000-6-2 + corr. September	2005 2005

¹⁾ Undated reference.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61326-1	2005	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements	EN 61326-1	2006
IEC 61326-2-1	2005	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-1: Particular requirements - Test configurations, operational conditions and performance criteria for sensitive test and measurement equipment for EMC unprotected applications	EN 61326-2-1	2006
IEC 61326-2-2	2005	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-2: Particular requirements - Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems	EN 61326-2-2	2006
IEC 61326-2-3	2006	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-3: Particular requirements - Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning	EN 61326-2-3	2006
IEC 61326-2-4	2006	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-4: Particular requirements - Test configurations, operational conditions and performance criteria for insulation monitoring devices according to IEC 61557-8 and for equipment for insulation fault location according to IEC 61557-9	EN 61326-2-4	2006
IEC 61326-2-5	2006	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-5: Particular requirements - Test configurations, operational conditions and performance criteria for field devices with interfaces according to IEC 61784-1, CP 3/2	EN 61326-2-5	2006
IEC 61326-3-1	2008	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 3-1: Immunity requirements for safety-related systems and for equipment intended to perform safety-related functions (functional safety) - General industrial applications	EN 61326-3-1	2008
IEC 61508-2	2000	Functional safety of electrical/electronic/programmable electronic safety-related systems - Part 2: Requirements for electrical/electronic/programmable electronic safety-related systems	EN 61508-2	2001

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61508-4 + corr. April	1998 1999	Functional safety of electrical/electronic/programmable electronic safety-related systems - Part 4: Definitions and abbreviations	EN 61508-4	2001
IEC 61511-1 + corr. November	2003 2004	Functional safety - Safety instrumented systems for the process industry sector - Part 1: Framework, definitions, system, hardware and software requirements	EN 61511-1	2004
ISO/IEC Guide 51	1999	Safety aspects - Guidelines for their inclusion - in standards		-

CONTENTS

INTRODUCTION.....	6
1 Scope.....	9
2 Normative references	10
3 Terms and definitions	11
4 General	13
5 EMC test plan.....	14
5.1 General.....	14
5.2 Configuration of EUT during testing.....	14
5.2.1 General	14
5.2.2 Composition of EUT.....	14
5.2.3 Assembly of EUT	14
5.2.4 I/O ports	15
5.2.5 Auxiliary equipment	15
5.2.6 Cabling and earthing (grounding).....	15
5.3 Operation conditions of EUT during testing.....	15
5.3.1 Operation modes	15
5.3.2 Environmental conditions.....	15
5.3.3 EUT software during test	15
5.4 Specification of performance criteria	15
5.5 Test description.....	15
6 Performance criteria	16
6.1 Performance criteria A, B and C	16
6.2 Performance criterion FS.....	16
6.3 Application of the performance criterion FS	16
7 Immunity requirements	16
8 Test set-up and test philosophy for EUT with functions intended for safety applications	19
8.1 Testing of safety-related systems and equipment intended to be used in safety-related systems	19
8.2 Test philosophy for equipment intended for use in safety-related systems.....	20
8.3 Test philosophy for safety-related systems	20
8.4 Test configuration.....	20
8.5 Monitoring	21
9 Test results and test report.....	21
Annex A (informative) Evaluation of electromagnetic phenomena	25
Annex B (informative) Details of the specified electromagnetic environment	27
Annex C (informative) Example of immunity levels in the process industry	31
Bibliography.....	32
Figure 1 – Correlation between the standards IEC 61326-1, IEC 61326-2-X, IEC 61326-3-1 and IEC 61326-3-2	8

Figure 2 – Typical test set-up for equipment intended for use in a safety related system integrated into a representative safety-related system during test	22
Figure 3 – Typical test set-up for equipment intended for use in a safety-related system tested stand-alone	23
Figure 4 – Test set-up for a safety-related system	24
Figure B.1 – Recommended cable layouts for different categories	28
Table 1a – Immunity test requirements for equipment intended for use in industrial locations with specified electromagnetic environment – Enclosure port	17
Table 1b – Immunity test requirements for equipment intended for use in industrial locations with specified electromagnetic environment – Input and output a.c. power ports	17
Table 1c – Immunity test requirements for equipment intended for use in industrial locations with specified electromagnetic environment – Input and output d.c. power ports	18
Table 1d – Immunity test requirements for equipment intended for use in industrial locations with specified electromagnetic environment – I/O signal/control ports	18
Table 1e – Immunity test requirements for equipment intended for use in industrial locations with specified electromagnetic environment – I/O signal/control ports connected directly to power supply networks	19
Table 1f – Immunity test requirements for equipment intended for use in industrial locations with specified electromagnetic environment – Functional earth port	19
Table A.1 – General considerations for the application of electromagnetic phenomena for functional safety in industrial applications with specified electromagnetic environment (examples)	26
Table C.1 – Immunity test requirements for equipment intended for use in industrial locations with a specified electromagnetic environment according to NE 21	31

INTRODUCTION

Functional safety is that part of the overall safety relating to the equipment under control (EUC) and the EUC control system which depends on the correct functioning of the electrical safety-related systems. To achieve this all items of equipment of the safety-related system which are involved in the performance of the safety functions must behave in a specified manner under all relevant conditions.

The IEC basic safety publication for functional safety of electrical/electronic/programmable electronic safety-related systems is IEC 61508. It sets the overall requirements to achieve functional safety. Sufficient immunity to electromagnetic disturbances is one of those requirements.

The concept of IEC 61508 distinguishes between the consideration of the application and the design of safety-related electrical and electronic systems. The interface between both is the safety requirements specification (SRS). It specifies all relevant requirements of the intended application, as follows.

- a) Definition of the safety function, based on a risk assessment of the intended application (which function is intended to reduce risk).
- b) Appropriate safety integrated level (SIL) for each safety-function based on a risk assessment of the intended application.
- c) Definition of the environment in which the system is intended to work including the electromagnetic environment as required by IEC 61508-2.

Hence, with regard to immunity against electromagnetic phenomena, the essential starting point is that the electromagnetic environment and its phenomena are considered in the SRS, as required by IEC 61508. The safety-related system intended to implement the specified safety-function has to fulfil the SRS, and from it corresponding immunity requirements have to be derived for the items of equipment; this results in an equipment requirement specification. With respect to the electromagnetic environment, the SRS and the equipment requirement specification should be based on a competent assessment of the foreseeable electromagnetic threats in the real environment over the whole operational life of the equipment. Hence immunity requirements for the equipment depend on the characteristics of the electromagnetic environment in which the equipment is intended to be used.

The equipment manufacturer, therefore, has to prove that the equipment fulfils the equipment requirement specification and the system integrator must prove that the system fulfils the SRS. Evidence has to be produced by application of appropriate methods. They do not need to consider any other aspects of the application, for example, risk of the application associated to any failure of the safety-related system. The objective is for all equipment in the system to comply with particular performance criteria taking into account functional safety aspects (for example the performance criterion FS) up to levels specified in the SRS independent of the required safety integrity level (SIL).

There are basically two approaches on how to deal with the electromagnetic environments and to conclude on immunity requirements.

- (A) To consider a general electromagnetic environment with no specific restrictions, for example an industrial environment, and to take into account all the electromagnetic phenomena that can occur as well as their maximum amplitudes when deriving appropriate immunity levels for the system and the equipment. This approach has been used to determine the levels specified within IEC 61326-3-1 leading to increased immunity levels for some electromagnetic phenomena compared to immunity levels which are derived without functional safety considerations.
- (B) To control the electromagnetic environment for example by the application of particular installation and mitigation practices, in such a way that electromagnetic phenomena and their amplitudes could occur only to a certain extent. These phenomena and restricted amplitudes are then taken into account by appropriate

immunity levels. These levels are not necessarily higher than those derived without functional safety considerations because it is ensured by corresponding means that no higher amplitudes as normally are to be expected. This approach is considered in this part of IEC 61326..

Applying approach (B) results in the fact that there is a specified electromagnetic environment due to the strict observation of particular installation and mitigation practices. In addition, however, appropriate knowledge is required concerning the electromagnetic phenomena and the amplitudes to be expected in this specified electromagnetic environment. This has been achieved by taking into account statistical data on faults in safety applications of the process industry. For this evaluation more than 20 000 units in safety applications are annually analysed on the occurrence of failures; from this data it has been shown that the failure rates meet the requirements connected to the safety integrity level (SIL). These units are in compliance with particular EMC requirements of the process industry.


Following approach (B), IEC 61326-3-2 gives specific electromagnetic immunity requirements that apply to safety-related systems and equipment intended to be used in safety-related systems. These requirements supplement some requirements of IEC 61326-1 (or of comparable EMC requirements of the process industry) and the selected electromagnetic phenomena and defined immunity test levels are expected to match with the environmental conditions of the specified industrial applications as defined in the scope of this standard.

The correlation between the standards IEC 61326-1, IEC 61326-2-X, IEC 61326-3-1 and IEC 61326-3-2 is described in the diagram of Figure 1.

The specified test levels in this standard are derived from the highest levels to be expected in the specified environment of industrial applications. These test levels are related to the electromagnetic environment (what can occur). They cannot be related in an analytical way to the SIL required for the safety-related system because there is no practically provable relationship between test level and probability of failure during use. The influences of electromagnetic phenomena are considered as systematic effects and by their nature often result in common cause events.

Design features of equipment must take into account the required SIL and must be designed to avoid dangerous systematic failures. Sufficient immunity against electromagnetic disturbances can only be ensured by design, mitigation and construction techniques which take into account electromagnetic aspects, which, however, are not within the scope of this standard.

It is therefore recommended that the approach to achieve the capability for the required SIL should be through the adoption of design features on the one hand and through appropriate test performance parameters in order to increase the level of confidence in the test results on the other hand.



NOTE 1 Functional safety requirements for equipment considered as proven in use according to IEC 61511-1 are excluded from the scope of IEC 61326-3-1 and 61326-3-2.
 NOTE 2 The term EMC tests refers to test levels of the relevant standards, for example, IEC 61326-1, IEC 61326-2-X or IEC 61000-6-2.
 NOTE 3 This flowchart does not intend to give requirements about the sequence of test.
 NOTE 4 For equipment intended to be used in safety related applications, see additional requirements in Table 4 of IEC 61326-3-1.

IEC 2338/07

Figure 1 – Correlation between the standards IEC 61326-1, IEC 61326-2-X, IEC 61326-3-1 and IEC 61326-3-2

ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE – EMC REQUIREMENTS –

Part 3-2: Immunity requirements for safety-related systems and for equipment intended to perform safety-related functions (functional safety) – Industrial applications with specified electromagnetic environment

1 Scope

The scope of IEC 61326-1 applies to this part of IEC 61326 but is limited to systems and equipment for industrial applications within a specified electromagnetic environment and intended to perform safety functions as defined in IEC 61508 with SIL 1-3.

The electromagnetic environments encompassed by this product family standard are industrial, both indoor and outdoor, as they can be found in industrial applications with an electromagnetic environment having specified characteristics (for example, process industry). The difference between the electromagnetic environment covered by this standard compared to the general industrial environment (see IEC 61326-3-1) is due to the mitigation measures employed against electromagnetic phenomena leading to a specified electromagnetic environment.

The environment of industrial application with a specified electromagnetic environment typically includes the following characteristics:

- industrial area with limited access;
- limited use of mobile transmitter;
- dedicated cables for power supply and control, signal or communication lines;
- separation between power supply and control, signal or communication cables;
- factory building mostly consisting of metal construction;
- overvoltage/lightning protection by appropriate measures (for example, metal construction of the building or use of protection devices);
- pipe heating systems driven by a.c. main power may be present;
- no high-voltage substation close to sensitive areas;
- presence of CISPR 11 Group 2 ISM equipment using ISM frequencies only with low power;
- competent staff;
- periodical maintenance of equipment and systems;
- mounting and installation guidelines for equipment and systems.

A more detailed description of the above-mentioned typical characteristics is given in Annex B.

Equipment and systems considered as “proven-in-use” according to IEC 61508 or IEC 61511 are excluded from the scope of IEC 61326-3-2.

Fire alarm systems and security alarm systems intended for protection of buildings are excluded from the scope of IEC 61326-3-2.