

© Copyright SEK Svensk Elstandard. Reproduction in any form without permission is prohibited.

REDLINE VERSION

Elektrisk utrustning för medicinskt bruk – Säkerhet –

Del 2-37: Särskilda fordringar på ultraljudsutrustning för diagnos och övervakning

Medical electrical equipment –

*Part 2-37: Particular requirements for the basic safety and essential performance of
ultrasonic medical diagnostic and monitoring equipment*

En så kallad ”Redline version” (RLV) innehåller både standarden som fastställts som SEK-publication och en ändringsmarkerad IEC-standard. Alla tillägg och borttagningar sedan den tidigare utgåvan av IEC-standarderna är markerade med färg. Med en RLV sparar du mycket tid när du ska identifiera och bedöma aktuella ändringar i standarden. SEK Svensk Elstandard kan bara ge ut RLV i de fall den finns tillgänglig från IEC.



IEC 60601-2-37

Edition 3.0 2024-07
REDLINE VERSION

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-37: Particular requirements for the basic safety and essential performance
of ultrasonic medical diagnostic and monitoring equipment**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.55; 17.140.50

ISBN 978-2-8322-9372-0

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	2
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	10
201.4 General requirements	17
201.5 General requirements for testing ME EQUIPMENT	18
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	18
201.7 ME EQUIPMENT identification, marking and documents	18
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	23
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	23
201.10 Protection against unwanted and excessive radiation HAZARDS	23
201.11 Protection against excessive temperatures and other HAZARDS	24
201.12 Accuracy of controls and instruments and protection against hazardous outputs	30
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	32
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	32
201.15 Construction of ME EQUIPMENT	32
201.16 ME SYSTEMS	32
201.17 *Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	32
202 ELECTROMAGNETIC DISTURBANCES – Requirements and tests.....	33
212 Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment (EMS)	36
Annexes	37
Annex AA (informative) Particular guidance and rationale	38
Annex BB (informative) Guidance in classification according to CISPR 11	47
Annex CC (informative) Guidance to the MANUFACTURER on the interpretation of <i>T_I</i> and <i>M_I</i> to be used to inform the OPERATOR.....	48
Annex DD (informative) Example set-up to measure surface temperature of externally applied TRANSDUCER ASSEMBLIES	52
Annex EE (informative) Acoustic output table intended for third parties.....	55
Bibliography.....	58
Index of defined terms	62
Figure AA.1 – Method a) for an external probe.....	41
Figure AA.2 – Method b) for an external probe.....	42
Figure AA.3 – Method b) for an external probe.....	42
Figure DD.1 – Set-up of an example test object to measure the surface temperature of externally applied transducers	54
Table 201.101 – List of symbols.....	16
Table 201.102 – Distributed essential performance requirements	18

Table 201.103 – Acoustic output reporting table	22
Table 201.104 – Overview of the tests noted under 201.11.1.3	29
Table CC.1 – Relative importance of maintaining low exposure indices in various scanning situations	50
Table DD.1 – Acoustic and thermal properties of tissues and materials.....	52
Table DD.2 – Weight % pure components	53
Table EE.1 – Example of acoustic output table for third parties	56

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-37: Particular requirements for the basic safety
and essential performance of ultrasonic medical
diagnostic and monitoring equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition IEC 60601-2-37:2007+AMD1:2015 CSV. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

IEC 60601-2-37 has been prepared by subcommittee 62B: Medical imaging equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This third edition cancels and replaces the second edition published in 2007 and Amendment 1:2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) technical and editorial changes resulting from the amended general standard IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and its collateral standards IEC 60601-1-xx,
- b) technical and editorial changes as a result of maintenance to normative references;
- c) technical and editorial changes resulting from relevant developments in TC 87 Ultrasonics standards. In particular, Clause 201.11 about protection against excessive temperatures and other hazards has been fully revised.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62B/1318/CDV	62B/1348/RVC

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

In this document, safety requirements additional to those in ~~the general standard IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020~~ are specified for ULTRASONIC DIAGNOSTIC EQUIPMENT.

A general guidance and rationale for the requirements of this document are given in Annex AA.

Knowledge of the reasons for these requirements will not only facilitate the proper application of this document but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology.

The approach and philosophy used in drafting this document for safety of ULTRASONIC DIAGNOSTIC EQUIPMENT are consistent with those in standards of the IEC 60601-2 series that apply to other diagnostic modalities, such as X-ray equipment and magnetic resonance systems.

In each case, the safety standard is intended to require increasing sophistication of output display indicators and controls with increasing energy levels in the interrogating field of diagnosis. Thus, for all such diagnostic modalities, it is the responsibility of the OPERATOR to understand the risk of the output of the ULTRASONIC DIAGNOSTIC EQUIPMENT, and to act appropriately in order to obtain the needed diagnostic information with the minimum risk to the PATIENT.

~~INTRODUCTION TO AMENDMENT 1~~

~~The second edition of IEC 60601-2-37 was published in 2007. Since that publication, the parent standard, IEC 60601-1:2005, entered maintenance, under which an amendment (IEC 60601-1:2005/AMD1:2012) and a consolidated edition 3.1 (IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012) were published. This amendment to IEC 60601-2-37:2007 addresses three issues:~~

- ~~1) technical changes proposed by National Committees as a result of 4 years of practical usage,~~
- ~~2) technical and editorial changes resulting from the amended general standard IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 and its collateral standards IEC 60601-1-xx, and~~
- ~~3) technical changes as a result of maintenance to normative references.~~

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment

~~The clauses and subclauses of the general standard apply except as follows:~~

201.1 Scope, object and related standards

Clause 1 of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 *Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ULTRASONIC DIAGNOSTIC EQUIPMENT as defined in 201.3.217, hereinafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in ~~7.2.13 and 8.4.1 of this standard~~ 201.7.2.13.

~~NOTE See also subclause 4.2 of this standard.~~

This document does not cover ultrasonic therapeutic equipment. Equipment used for the imaging or diagnosis of body structures by ultrasound in conjunction with other medical procedures is covered.

201.1.2 Object

Replacement:

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ULTRASONIC DIAGNOSTIC EQUIPMENT as defined in 201.3.217.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and Clause 201.2 of this document.

IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 apply as modified in Clause 202 and Clause 212 respectively. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

~~For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.~~

The numbering of sections, clauses and subclauses of this document corresponds to that of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.6 in this document addresses the content of Clause 6 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is replaced completely by this document.

"*Addition*" means that the text of this document is additional to the requirements of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard.

"*Amendment*" means that the clause or subclause of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the ~~general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, any applicable collateral standards and this document taken together.

Where there is no corresponding section, clause or subclause in this document, the section, clause or subclause of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or

applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

Clause 2 of ~~the general standard~~ IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies except as follows:

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012¹

IEC 60601-1:2005/AMD2:2020

IEC 60601-1-12:2014, *Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

IEC 60601-1-12:2014/AMD1:2020

IEC 60601-2-18:2009, *Medical electrical equipment – Part 2-18: Particular requirements for the basic safety and essential performance of endoscopic equipment*

IEC 62127-1:~~2007~~2022, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields up to 40 MHz*

~~IEC 62127-1:2007/AMD1:2013~~²

IEC 62359:2010, *Ultrasonics – Field characterization – Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields*

IEC 62359:2010/AMD1:2017

CISPR 11:2024, *Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement*

¹~~There exists a consolidated edition (3.1) including IEC 60601-1:2005 and its Amendment 1 (2012).~~

²~~There exists a consolidated edition (1.1) including IEC 62127-1:2007 and its Amendment 1 (2013).~~

© Copyright SEK Svensk Elstandard. Reproduction in any form without permission is prohibited.

Elektrisk utrustning för medicinskt bruk – Säkerhet – Del 2-37: Särskilda fordringar på ultraljudsutrustning för diagnos och övervakning

Medical electrical equipment –

*Part 2-37: Particular requirements for the basic safety and essential performance of
ultrasonic medical diagnostic and monitoring equipment*

Som svensk standard gäller europastandarden EN IEC 60601-2-37:2024. Den svenska standarden innehåller den officiella engelska språkversionen av EN IEC 60601-2-37:2024.

Nationellt förord

Europastandarden EN IEC 60601-2-37:2024

består av:

- **europastandardens ikraftsättningsdokument**, utarbetat inom CENELEC
- **IEC 60601-2-37, Third edition, 2024 - Medical electrical equipment - Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment**

utarbetad inom International Electrotechnical Commission, IEC.

Tidigare fastställd svensk standard SS-EN 60601-2-37, utg 2:2008 med eventuella tillägg, ändringar och rättelser gäller ej fr o m 2027-11-30.

ICS 11.040.55; 17.140.50

Denna standard är fastställd av SEK Svensk Elstandard,
som också kan lämna upplysningar om **sakinnehållet** i standarden.
Postadress: Box 1042, 172 21 Sundbyberg
Telefon: 08 - 444 14 00.
E-post: sek@elstandard.se. Internet: elstandard.se

Standarder underlättar utvecklingen och höjer elsäkerheten

Det finns många fördelar med att ha gemensamma tekniska regler för bl a mätning, säkerhet och provning och för utförande, skötsel och dokumentation av elprodukter och elanläggningar.

Genom att utforma sådana standarder blir säkerhetsfordringar tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

SEK är Sveriges röst i standardiseringsarbetet inom elområdet

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

Stora delar av arbetet sker internationellt

Utformningen av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringsarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringsverksamhet och medlemsavgift till IEC och CENELEC.

Var med och påverka!

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtida standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

SEK Svensk Elstandard

Box 1042
172 21 Sundbyberg
Tel 08-444 14 00
elstandard.se

English Version

Medical electrical equipment - Part 2-37: Particular requirements
for the basic safety and essential performance of ultrasonic
medical diagnostic and monitoring equipment
(IEC 60601-2-37:2024)

Appareils électromédicaux - Partie 2-37: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils de diagnostic et de surveillance
médicaux à ultrasons
(IEC 60601-2-37:2024)

Medizinische elektrische Geräte - Teil 2-37: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Ultraschallgeräten für
die medizinische Diagnose und Überwachung
(IEC 60601-2-37:2024)

This European Standard was approved by CENELEC on 2024-10-16. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62B/1318/CDV, future edition 3 of IEC 60601-2-37, prepared by SC 62B "Medical imaging equipment, software, and systems" of IEC/TC 62 "Medical equipment, software, and systems" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-37:2024.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2025-11-30 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2027-11-30 document have to be withdrawn

This document supersedes EN 60601-2-37:2008 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 60601-2-37:2024 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 61157:2007	NOTE	Approved as EN 61157:2007 (not modified)
IEC 61157:2007/A1:2013	NOTE	Approved as EN 61157:2007/A1:2013 (not modified)
IEC 60601-1-11:2015	NOTE	Approved as EN 60601-1-11:2015 (not modified)
IEC 60601-1-11:2015/A1:2020	NOTE	Approved as EN 60601-1-11:2015/A1:2021 (not modified)

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Clause 2 of EN 60601-1:2006, EN 60601-1:2006/AC:2010, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A1:2013/AC:2014, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021, EN 60601-1:2006/AC:2022 and EN 60601-1:2006/A13:2024 is applicable, except as follows:

Add the following references:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-		+ AC	2010
+ A1	2012		+ A1	2013
-	-		+ AC	2014
-	-		+ A12	2014
+ A2	2020		+ A2	2021
-	-		+ AC	2022
-	-		+ A13	2024
IEC 60601-1-12	2014	Medical electrical equipment - Part 1-12: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment	EN 60601-1-12	2015
+ A1	2020		+ A1	2020
IEC 60601-2-18	2009	Medical electrical equipment - Part 2-18: Particular requirements for the basic safety and essential performance of endoscopic equipment	EN 60601-2-18	2015
IEC 62127-1	2022	Ultrasonics - Hydrophones - Part 1: Measurement and characterization of medical ultrasonic fields	EN IEC 62127-1	2022

EN IEC 60601-2-37:2024 (E)

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 62359	2010	Ultrasonics - Field characterization - Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields	EN 62359	2011
+ A1	2017		+ A1	2018
CISPR 11	2024	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement	EN 55011	— ¹

¹ Under preparation. Stage at the time of publication: FprEN 55011:2023.

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Medical electrical equipment –

Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment

Appareils électromédicaux –

Partie 2-37: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de diagnostic et de surveillance médicaux à ultrasons

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.55, 17.140.50

ISBN 978-2-8322-8920-4

<p>Warning! Make sure that you obtained this publication from an authorized distributor.</p> <p>Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.</p>
--

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	10
201.4 General requirements	17
201.5 General requirements for testing ME EQUIPMENT	18
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	18
201.7 ME EQUIPMENT identification, marking and documents	18
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	23
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	23
201.10 Protection against unwanted and excessive radiation HAZARDS	23
201.11 Protection against excessive temperatures and other HAZARDS	24
201.12 Accuracy of controls and instruments and protection against hazardous outputs	29
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	31
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	31
201.15 Construction of ME EQUIPMENT	31
201.16 ME SYSTEMS	31
201.17 *Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	31
202 ELECTROMAGNETIC DISTURBANCES – Requirements and tests.....	31
212 Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment (EMS)	33
Annexes	34
Annex AA (informative) Particular guidance and rationale.....	35
Annex BB (informative) Guidance in classification according to CISPR 11	43
Annex CC (informative) Guidance to the MANUFACTURER on the interpretation of <i>T_I</i> and <i>M_I</i> to be used to inform the OPERATOR.....	44
Annex DD (informative) Example set-up to measure surface temperature of externally applied TRANSDUCER ASSEMBLIES	48
Annex EE (informative) Acoustic output table intended for third parties.....	51
Bibliography.....	54
Index of defined terms	58
Figure AA.1 – Method a) for an external probe.....	38
Figure AA.2 – Method b) for an external probe.....	39
Figure AA.3 – Method b) for an external probe.....	39
Figure DD.1 – Set-up of an example test object to measure the surface temperature of externally applied transducers	50
Table 201.101 – List of symbols.....	16
Table 201.102 – Distributed essential performance requirements	18

Table 201.103 – Acoustic output reporting table	22
Table 201.104 – Overview of the tests noted under 201.11.1.3	28
Table CC.1 – Relative importance of maintaining low exposure indices in various scanning situations	46
Table DD.1 – Acoustic and thermal properties of tissues and materials.....	48
Table DD.2 – Weight % pure components	49
Table EE.1 – Example of acoustic output table for third parties	52

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-37: Particular requirements for the basic safety
and essential performance of ultrasonic medical
diagnostic and monitoring equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-37 has been prepared by subcommittee 62B: Medical imaging equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This third edition cancels and replaces the second edition published in 2007 and Amendment 1:2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) technical and editorial changes resulting from the amended general standard IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and its collateral standards IEC 60601-1-xx,

- b) technical and editorial changes as a result of maintenance to normative references;
- c) technical and editorial changes resulting from relevant developments in TC 87 Ultrasonics standards. In particular, Clause 201.11 about protection against excessive temperatures and other hazards has been fully revised.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62B/1318/CDV	62B/1348/RVC

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

In this document, safety requirements additional to those in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 are specified for ULTRASONIC DIAGNOSTIC EQUIPMENT.

A general guidance and rationale for the requirements of this document are given in Annex AA.

Knowledge of the reasons for these requirements will not only facilitate the proper application of this document but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology.

The approach and philosophy used in drafting this document for safety of ULTRASONIC DIAGNOSTIC EQUIPMENT are consistent with those in standards of the IEC 60601-2 series that apply to other diagnostic modalities, such as X-ray equipment and magnetic resonance systems.

In each case, the safety standard is intended to require increasing sophistication of output display indicators and controls with increasing energy levels in the interrogating field of diagnosis. Thus, for all such diagnostic modalities, it is the responsibility of the OPERATOR to understand the risk of the output of the ULTRASONIC DIAGNOSTIC EQUIPMENT, and to act appropriately in order to obtain the needed diagnostic information with the minimum risk to the PATIENT.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment

201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 *Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ULTRASONIC DIAGNOSTIC EQUIPMENT as defined in 201.3.217, hereinafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 201.7.2.13.

This document does not cover ultrasonic therapeutic equipment. Equipment used for the imaging or diagnosis of body structures by ultrasound in conjunction with other medical procedures is covered.

201.1.2 Object

Replacement:

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ULTRASONIC DIAGNOSTIC EQUIPMENT as defined in 201.3.217.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and Clause 201.2 of this document.

IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 apply as modified in Clause 202 and Clause 212 respectively. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

The numbering of sections, clauses and subclauses of this document corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.6 in this document addresses the content of Clause 6 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

"Replacement" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is replaced completely by this document.

"Addition" means that the text of this document is additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard.

"Amendment" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, any applicable collateral standards and this document taken together.

Where there is no corresponding section, clause or subclause in this document, the section, clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies except as follows:

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC 60601-1-12:2014, *Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

IEC 60601-1-12:2014/AMD1:2020

IEC 60601-2-18:2009, *Medical electrical equipment – Part 2-18: Particular requirements for the basic safety and essential performance of endoscopic equipment*

IEC 62127-1:2022, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields*

IEC 62359:2010, *Ultrasonics – Field characterization – Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields*

IEC 62359:2010/AMD1:2017

CISPR 11:2024, *Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement*