

INTERNATIONAL STANDARD

IEC
61010-2-061

Second edition
2003-06

GROUP SAFETY PUBLICATION

Safety requirements for electrical equipment for measurement, control, and laboratory use –

Part 2-061: Particular requirements for laboratory atomic spectrometers with thermal atomization and ionization

*Règles de sécurité pour appareils électriques
de mesure, de régulation et de laboratoire –*

*Partie 2-061:
Prescriptions particulières pour spectromètres de
laboratoire avec vaporisation et ionisation thermiques*

© IEC 2003 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

N

For price, see current catalogue

CONTENTS

FOREWORD	3
1 Scope and object	5
2 Normative references	5
3 Definitions	5
4 Tests	6
5 Marking and documentation	6
6 Protection against electric shock	9
7 Protection against mechanical hazards	9
8 Mechanical resistance to shock and impact	9
9 Protection against the spread of fire	9
10 Equipment temperature limits and resistance to heat	9
10.1 Surface temperature limits for protection against burns	9
11 Protection against HAZARDS from fluids	9
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	10
13 Protection against liberated gases, explosion and implosion	11
14 Components	12
15 Protection by interlocks	13
16 Test and measurement equipment	13
Annexes	14
Annex F (normative) ROUTINE TESTS	14
Annex H (informative) Index of defined terms	14
Bibliography	14

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT
FOR MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 2-061: Particular requirements for laboratory atomic
spectrometers with thermal atomization and ionization**

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-061 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

This second edition cancels and replaces the first edition published in 1995, of which it constitutes a technical revision.

It has the status of a group safety publication in accordance with IEC Guide 104.

The text of this standard is based on the following documents:

FDIS	Report on voting
66/326/FDIS	66/331/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

This Part 2-061 is intended to be used in conjunction with IEC 61010-1. It was established on the basis of the second edition (2001). Consideration may be given to future editions of, or amendments to, IEC 61010-1.

This Part 2-061 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Safety requirements for laboratory atomic spectrometers with thermal atomization and ionization*.

Where a particular subclause of Part 1 is not mentioned in this part 2, that subclause applies as far as is reasonable. Where this part states “addition”, “modification” or “replacement”, or “deletion”, the relevant requirement, test specification or note in Part 1 should be adapted accordingly.

In this standard:

- 1) the following print types are used:
 - requirements: in roman type;
 - NOTES: in small roman type;
 - *conformity and test: in italic type*;
 - terms used throughout this standard which have been defined in clause 3: SMALL ROMAN CAPITALS;
- 2) subclauses, figures, tables and notes which are additional to those in Part 1 are numbered starting from 101.

The committee has decided that the contents of this publication will remain unchanged until 2007. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 2-061: Particular requirements for laboratory atomic spectrometers with thermal atomization and ionization

1 Scope and object

This clause of Part 1 is applicable except as follows:

1.1 Scope

1.1.1 Equipment included in scope

Replacement:

This part of IEC 61010 applies to electrically powered laboratory atomic spectrometers with thermal atomization.

NOTE 1 Examples include atomic absorption spectrometers, emission flame photometers, atomic fluorescence spectrophotometers, inductively coupled plasma spectrometers, microwave coupled plasma spectrometers and mass spectrometers, all with thermal atomization and ionization (including tubing and connectors which are provided by the manufacturer for connection to external supplies).

NOTE 2 If all or part of the equipment falls within the scope of one or more other part 2 standards of IEC 61010 as well as within the scope of this standard, it will also need to meet the requirements of those other part 2 standards.