

© Copyright SEK. Reproduction in any form without permission is prohibited.

**Elektriska apparater för detektering och mätning av  
brännbara eller giftiga gaser i parkeringsgarage och tunnlar –  
Del 1: Allmänna prestandafordringar samt metoder för provning av  
detektering och mätning av kolmonoxid och kväveoxid**

*Electrical apparatus for the detection and measurement of toxic and combustible gases  
in car parks and tunnels –*

*Part 1: General performance requirements and test methods for the detection  
and measurement of carbon monoxide and nitrogen oxides*

Som svensk standard gäller europastandarden EN 50545-1:2011. Den svenska standarden innehåller den officiella engelska språkversionen av EN 50545-1:2011.

---

ICS 13.040.50

Denna standard är fastställd av SEK Svensk Elstandard,  
som också kan lämna upplysningar om **sakinnehållet** i standarden.  
Postadress: SEK, Box 1284, 164 29 KISTA  
Telefon: 08 - 444 14 00. Telefax: 08 - 444 14 30  
E-post: sek@elstandard.se. Internet: www.elstandard.se

---

## *Standarder underlättar utvecklingen och höjer elsäkerheten*

Det finns många fördelar med att ha gemensamma tekniska regler för bl a säkerhet, prestanda, dokumentation, utförande och skötsel av elprodukter, elanläggningar och metoder. Genom att utforma sådana standarder blir säkerhetskraven tydliga och utvecklingskostnaderna rimliga samtidigt som marknadens acceptans för produkten eller tjänsten ökar.

Många standarder inom elområdet beskriver tekniska lösningar och metoder som åstadkommer den elsäkerhet som föreskrivs av svenska myndigheter och av EU.

## *SEK är Sveriges röst i standardiseringssarbetet inom elområdet*

SEK Svensk Elstandard svarar för standardiseringen inom elområdet i Sverige och samordnar svensk medverkan i internationell och europeisk standardisering. SEK är en ideell organisation med frivilligt deltagande från svenska myndigheter, företag och organisationer som vill medverka till och påverka utformningen av tekniska regler inom elektrotekniken.

SEK samordnar svenska intressenters medverkan i SEKs tekniska kommittéer och stödjer svenska experters medverkan i internationella och europeiska projekt.

## *Stora delar av arbetet sker internationellt*

Utdriften av standarder sker i allt väsentligt i internationellt och europeiskt samarbete. SEK är svensk nationalkommitté av International Electrotechnical Commission (IEC) och Comité Européen de Normalisation Electrotechnique (CENELEC).

Standardiseringssarbetet inom SEK är organiserat i referensgrupper bestående av ett antal tekniska kommittéer som speglar hur arbetet inom IEC och CENELEC är organiserat.

Arbetet i de tekniska kommittéerna är öppet för alla svenska organisationer, företag, institutioner, myndigheter och statliga verk. Den årliga avgiften för deltagandet och intäkter från försäljning finansierar SEKs standardiseringssverksamhet och medlemsavgift till IEC och CENELEC.

## *Var med och påverka!*

Den som deltar i SEKs tekniska kommittéarbete har möjlighet att påverka framtidens standarder och får tidig tillgång till information och dokumentation om utvecklingen inom sitt teknikområde. Arbetet och kontakterna med kollegor, kunder och konkurrenter kan gynnsamt påverka enskilda företags affärsutveckling och bidrar till deltagarnas egen kompetensutveckling.

Du som vill dra nytta av dessa möjligheter är välkommen att kontakta SEKs kansli för mer information.

### **SEK Svensk Elstandard**

Box 1284  
164 29 Kista  
Tel 08-444 14 00  
[www.elstandard.se](http://www.elstandard.se)

English version

**Electrical apparatus for the detection and measurement of toxic and combustible gases in car parks and tunnels -  
Part 1: General performance requirements and test methods for the detection and measurement of carbon monoxide and nitrogen oxides**

Appareil électrique de détection de mesure de gaz combustible et toxique dans les parcs de stationnement et les tunnels – Partie 1: Exigences de performance générales et méthodes pour la détection et la mesure du monoxyde de carbone et d'oxyde d'azote

Elektrische Geräte für die Detektion und Messung von toxischen (und brennbaren) Gasen in Tiefgaragen und Tunneln - Teil 1: Allgemeine Anforderungen an das Betriebsverhalten sowie Prüfverfahren für die Detektion und Messung von Kohlenmonoxid und Stickoxiden

This European Standard was approved by CENELEC on 2011-08-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

**CENELEC**

European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**Management Centre: Avenue Marnix 17, B - 1000 Brussels**

## Contents

<b>Foreword .....</b>	<b>4</b>
<b>Introduction.....</b>	<b>5</b>
<b>1 Scope .....</b>	<b>6</b>
<b>2 Normative references .....</b>	<b>7</b>
<b>3 Terms and definitions .....</b>	<b>7</b>
3.1 Gas properties .....	7
3.2 Types of apparatus and components .....	8
3.3 Sensors.....	8
3.4 Signals and alarms .....	9
3.5 Time.....	9
3.6 Other.....	11
<b>4 General requirements.....</b>	<b>11</b>
4.1 Introduction .....	11
4.2 Construction requirements .....	11
4.3 Operational requirements .....	14
4.4 Access level.....	16
4.5 Information for the user .....	17
<b>5 Test requirements.....</b>	<b>18</b>
5.1 General .....	18
5.2 Sequence of tests .....	18
5.3 Preparation of apparatus before testing .....	19
5.4 Test gas application.....	19
5.5 Normal conditions for test.....	20
<b>6 Test methods.....</b>	<b>21</b>
6.1 General .....	21
6.2 Unpowered storage .....	21
6.3 Linearity .....	21
6.4 Alarm set points and outputs (car parks only).....	22
6.5 Repeatability .....	23
6.6 Temperature .....	23
6.7 Humidity.....	24
6.8 Air velocity (for diffusion apparatus only) .....	24
6.9 Flow rate (for aspirated apparatus only).....	24
6.10 Interfering gases.....	24
6.11 Recovery from high gas concentrations .....	25
6.12 Mechanical strength .....	25
6.13 Warm-up time .....	25
6.14 Response time.....	25
6.15 Power supply variations.....	26
6.16 Long-term stability .....	26
6.17 Field calibration, or field verification kit.....	26
6.18 Calibration mask.....	26
6.19 Testing of multiple point selector .....	26
6.20 Vibration.....	27
6.21 Electromagnetic compatibility .....	27
<b>Annex A (informative) Using NO and NO<sub>2</sub> standard test gases for standard and humidity testing .....</b>	<b>28</b>
A.1 General .....	28
A.2 NO <sub>2</sub> gas testing.....	28
A.3 Humidity testing .....	28
<b>Annex B (informative) Clarification regarding “6.4 Alarm set points and outputs (car parks only)” .....</b>	<b>29</b>
<b>Bibliography.....</b>	<b>31</b>

**Figures**

Figure 1 – Warm-up time in clean air (typical).....	10
Figure 2 – Warm-up time in standard test gas (typical) .....	10
Figure B.1 – Output of the sensor .....	29
Figure B.2 – Time-weighted measured gas concentration.....	29
Figure B.3 – Missing areas for TWA calculation .....	30

**Tables**

Table 1 – Measuring range.....	14
Table 2 – Alarm levels .....	14
Table 3 – Maximum uncertainty and concentration of standard test gases.....	20
Table 4 – Tolerance for linearity .....	22
Table 5 – ppm.min calculated from Table 2 and alarms 1,2 test gas concentrations .....	22
Table 6 – Tolerances for alarm 1 and alarm 2 activation .....	23
Table 7 – Tolerances for zero deviation .....	23
Table 8 – Interfering gases table: test results to be inserted into the blank cells.....	24
Table 9 – Gas concentration for recovery test .....	25

## Foreword

This document (EN 50545-1:2011) has been prepared by CLC Technical Body 216 "Gas detectors".

The following dates are fixed:

- latest date by which this document has (dop) 2012-08-15  
to be implemented at national level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2014-08-15

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

## Introduction

This European Standard does not give guidance on installation of a gas detection system.

This European Standard specifies unique ranges and alarm levels for type testing, specifying minimum requirements. All alarm levels are variable and may be adapted to national and/or local regulations.

Separate type testing of remote gas sensors and control units is permitted. It is common practice to use an integrated tunnel control system that includes processing of toxic gas measurement along with other monitoring functions. When remote gas detectors and control units are type tested separately, it is the responsibility of the manufacturer to ensure that the assembled system complies with the requirements of this European Standard. The maximum capacity of the assembled system should not exceed the capacity of the system that has been type tested.

Engineering companies or installers who buy equipment from different manufacturers are responsible for the proper integration of the system.

## 1 Scope

This European Standard applies to apparatus for the detection and/or the measurement of carbon monoxide (CO), nitrogen monoxide (NO) and nitrogen dioxide (NO<sub>2</sub>) intended to control a ventilation system and/or to give an indication, alarm or any other signal to warn of a toxic hazard. These three gases are generically called “target gases” for the purpose of this European Standard.

National and local regulations might not require detection of NO or NO<sub>2</sub> and might require detection of other gases or vapours.

This European Standard includes requirements for remote gas sensors (RGS) to be used in car parks and tunnels and requirements for the control unit (CU) to be used in car parks.

This European Standard specifies general requirements for construction and testing and describes the test methods that apply to fixed apparatus for the detection and/or the measurement of the concentration of the target gases in car parks and tunnels. This European Standard may also be applied to similar applications where the concentration of the target gases could lead to a risk to health, for example loading areas for trucks and underground bus stations.

This European Standard also applies when an apparatus manufacturer makes any claims regarding superior performance that exceeds these minimum requirements.

This European Standard applies to apparatus, including the sampling system if applicable.

This European Standard does not specify requirements for apparatus to be installed in hazardous areas.

This European Standard does not apply for applications already covered by the following standards:

- domestic premises, covered by EN 50291-1;
- boats, craft, caravans or mobile homes, covered by EN 50291-2;
- workplace atmospheres, covered by EN 45544 series;
- emissions of heaters, covered by EN 50379 series;
- motor vehicles emissions, covered by ISO/PAS 3930;
- monitoring of the LEL level of combustible gases, covered by EN 60079-29-1.

This European Standard does not apply for the following applications and technologies:

- confined spaces not accessible to people;
- laboratory or analytical equipment;
- apparatus used to control industrial processes;
- portable and transportable apparatus;
- open path gas detection;
- tunnel construction;
- monitoring of particulates and dust;
- monitoring of combustible gases;
- CO monitoring for fire detection.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 50270, *Electromagnetic compatibility – Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen*

EN 50271, *Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen – Requirements and tests for apparatus using software and/or digital technologies*

EN 60073, *Basic and safety principles for man-machine interface, marking and identification – Coding principles for indicators and actuators* (IEC 60073)

EN 60335-1:2002 + corr. Jul.2009 + corr. May.2010 + A1:2004 + corr. Jan.2007 + A2:2006 + A11:2004 + A12:2006 + corr. Feb.2007 + A13:2008 + A14:2010, *Household and similar electrical appliances – Safety – Part 1: General requirements* (IEC 60335-1:2001 + A1:2004 + A2:2006 + corrigendum Aug. 2006)

EN 60529, *Degrees of protection provided by enclosures (IP Code)* (IEC 60529)